

Radiochemistry Developments on $^{224}\text{Radium}$ (^{224}Ra)

MUHAMMAD INZAMAM

(PINSTECH, Pakistan Atomic Energy Commission)

Layout

- Introduction
- Production of ^{224}Ra (^{224}Ra)
- $^{224}\text{Ra}/^{212}\text{Pb}$ Generator Production
- Labelling of ^{212}Pb with DOTA
- Conclusion



TAT Radionuclides

$^{227}\text{Th}/^{223}\text{Ra}$

$^{225}\text{Ac}/^{213}\text{Bi}$

^{211}At

$^{212}\text{Pb}/^{212}\text{Bi}$

$^{230}\text{U}/^{226}\text{Th}$

^{149}Tb



Introduction

- ^{224}Ra Radium is a pure alpha emitter
- Favorable half life 3.6 day
- Short lived daughter product


Production of $^{224}\text{Radium}$

- Several methods for the production
 - Radiochemical separation from the Th/U
 - Spallation reaction of thorium by proton irradiation
 - $^{228}\text{Th}/^{224}\text{Ra}$ Generator




$^{224}\text{Ra}/^{212}\text{Pb}$ Lead Generator Production

- There are two methods of $^{224}\text{Ra}/^{212}\text{Pb}$ generator production
 - Collection of emanated radon from radium source
 - Radiochemical separation of using cation exchange resin.




$^{224}\text{Ra}/^{212}\text{Pb}$ Generator by Radon Collection

- ^{224}Ra Radium is used as source in the generator
- Glass bottle with plastic lid as generator body
- Source adsorbed on the glass wool
- Collection chamber for radon
- Decays into ^{212}Pb




Elution of Generator

- Elution of generator after the establishment of secular equilibrium
- Elution of ^{212}Pb with 0.1M HCl




Efficiency and Stability

- Generator from few kBq to 80MBq are produced
- Elution are taken every two days for several weeks
- Generators are eluted with average efficiency of 33%
- High activity did not damage the generator structure




$^{224}\text{Ra}/^{212}\text{Pb}$ Generator by Cation exchange resin

- Generator consist of a glass column
- Cation exchange resin for adsorption of radium source
- Preconditioning of column
- ^{224}Ra loading on the column followed by washing
- Elution of generator with few ml of HCl




Efficiency of the Generator

- Generator eluted for three time after every two day
- Average 56% of the ^{212}Pb recovered from the generator
- Ra breakthrough was observed in the elution up to 9%
- Some issues identified during experiment which will lead to improvement



Radiation Safety

- All the following steps for generators carried out in Gloves box
 - ^{224}Ra foil retrieval for target holder
 - Dissolution of the ^{224}Ra
 - Generator production
 - Elution
- No release of radon or any contamination



Quality of ^{212}Pb from Generator

- Gamma spec. analysis of elution
- In first generator, no radium was detected ($<^{224}\text{Ra}$ MDA) in elution
- Generator of activity $\sim 80\text{MBq}$ dispatch to one of partner institute for further evaluation/labelling studies
- In cation exchange resin, up to 9% radium was detected used internally only on experimental work



9/21/2023 5:15:53PM Page 1 of 7

Analysis Report for CORMED-Inzi-Ra-224-blue-bottle2
CORMED-Inzi-Ra-224-blue-bottle2-after-change-bouchon-250cm

HSE-RP RADIO-ANALYTICAL LAB - GAMMA SPECTROSCOPY ANALYSIS

Sample Number : 339319
Sample Identification : CORMED-Inzi-Ra-224-blue-bottle2
Sample Description : CORMED-Inzi-Ra-224-blue-bottle2-after-change-bouchon-250cm
Sample Type : CORMED
Sample Size : 1.000E+00 units
Facility : RP
Detector Name : MED01-B22158
Operator : Med Operator 1
Sample Taken / Activity reported On : 9/21/2023 4:06:05PM
Acquisition Started : 9/21/2023 4:06:05PM
Procedure : Acquisition
Live Time : 900.0 seconds
Real Time : 907.9 seconds
Dead Time : 0.87 %

Peak Locate Threshold : 3.00
Peak Locate Range (in channels) : 25 - 8192
Peak Area Range (in channels) : 25 - 8192
Identification Energy Tolerance : 1.000 keV
Geometry : GEO-CORMED
Energy Calibration Used Done On : 8/21/2023 9:31:54AM
Efficiency Calibration Used Done On : 9/21/2023 5:06:33PM
Efficiency Calibration Description : CORMED-Ra-224-Inzi-blue-bottle-250cm
Background File : \\rp-apex-sv19\apex_data\RP\Data\0000314237.CNF

ATTENTION, nous consulter pour éviter toute
erreur d'interprétation
Rapport détaillé / officiel fourni à la demande

Labelling and Further Studies

- Radon collection type generator of 80MBq
dispatch to partner institute for further studies
 - Labelling of the ^{212}Pb with DOTATATE
 - Stability of Pb-DOTA Complex in mouse serum
- Request through PRISMAP for clinical translation studies in Dresden University Hospital

Conclusion



Production of $^{224}\text{Ra}/^{212}\text{Pb}$ generator of two different types



With generator efficiencies up to 33% and 56%.



High quality product with safe operation



Good labelling result and stability results

*Thank
you*

