



Contribution ID: 84

Type: Poster

Concept Design of New Compact Electron Cyclotron Resonance Ion Source with Permanent Magnets for Multi-Ion Radiotherapy

The multi-ion radiotherapy with dose distribution and Liner Energy Transfer optimization is being studied at QST. Helium, carbon, oxygen and neon ions are considered as ion species for multi-ion therapy. Requirement values of beam current were 940 μA correspond to He^{2+} , 290 μA to C^{4+} , 330 μA to O^{6+} , and 245 μA to Ne^{7+} , respectively. We performed some beam tests for design of a new compact ion source with the existing 18 GHz NIRS-HEC. From the results of the beam tests, we estimate the mirror magnetic field of NIRS-HEC. Then, the structure of the permanent magnets for new ECR ion source is determined so that they reproduce the values of the upstream mirror peak (Binj), B minimum (Bmin) and the downstream mirror peak (Bext) at the NIRS-HEC. The magnetic field of Binj, Bmin and Bext at NIRS-HEC were 1.14 T, 0.475 T and 0.9 T, respectively. We will describe about the beam tests with the NIRS-HEC and design of the new compact ECR ion source.

E-mail for contact person

muramatsu.masayuki@qst.go.jp

Funding Information

Primary author: MURAMATSU, Masayuki (National Institutes for Quantum and Radiological Science and Technology)

Co-authors: Mr OUCHI, Fumihisa (Accelerator Engineering Corporation); SUZUKI, Taku (Accelerator Engineering Corporation); Mr TAKAHASHI, Katsuyuki (Accelerator Engineering Corporation); Mr SASANO, Toshinobu (Accelerator Engineering Corporation); Mr SHIRAISHI, Tadahiro (Accelerator Engineering Corporation); Dr NOMURA, Shinji (Sumitomo Heavy Industries); TAKAHASHI, NOBUAKI (Sumitomo Heavy Industries); KATO, Yushi (Osaka Univ.); Dr KATAGIRI, Ken (National Institutes for Quantum and Radiological Science and Technology); MIZUSHIMA, Kota (National Institutes for Quantum and Radiological Science and Technology); Dr IWATA, Yoshiyuki (National Institutes for Quantum and Radiological Science and Technology); Dr INANIWA, Taku (National Institutes for Quantum and Radiological Science and Technology); KITAGAWA, Atsushi (National Institutes for Quantum and Radiological Science and Te)

Presenter: MURAMATSU, Masayuki (National Institutes for Quantum and Radiological Science and Technology)

Session Classification: Poster Session 2

Track Classification: Applications and related technologies