



Contribution ID: 93

Type: **Poster**

Preliminary Results of Positive Ion Mass Spectrometry Based on a 2.45 GHz ECR Ion Source and a Non-Metallic Gas Target

Positive-ion mass spectrometry (PIMS) is a positive-to-negative ion conversion metrology instead of traditional accelerator mass spectrometry (AMS) on high-precision radiocarbon dating. To demonstrate the probability of PIMS with a 2.45 GHz ECR ion source, an experiment was launched at Peking University (PKU). A compact permanent magnet 2.45 GHz ECR ion source is designed to produce carbon ions from CO₂. A CXC unit was installed after the extraction system and ethylene is chosen as charge exchange gas. A dipole is employed for beam analysis. A faraday cup installed at the end of the beamline is used for beam current measurement. Preliminary experiments prove that more than hundreds of microamp of C²⁺ beam can be produced with this ECR source. By injecting ethylene into CXC, several microamp C⁻ beam was detected and the C²⁺ ions charge exchange efficiency to C⁻ ions can reach to 5%.

E-mail for contact person

sxpeng@pku.edu.cn

Funding Information

NSFC 11975036, 11775007

Primary authors: Dr ZHANG, Ailin (USTC); Mr CUI, Bujian (PKU); Dr SUN, Jiang (PKU); Mr LI, Kai (PKU); PENG, SHIXIANG (Peking University); Mr MA, Tenghao (PKU); Dr WU, Wenbin (PKU); Mr JIANG, Yaoxiang (PKU); Prof. GUO, Zhiyu (PKU)

Presenter: PENG, SHIXIANG (Peking University)

Session Classification: Poster Session 2

Track Classification: Applications and related technologies