



Canadian Association
of Physicists

Association canadienne
des physiciens et physiciennes

Contribution ID: 3509 Type: **Poster Competition (Graduate Student) / Compétition affiches (Étudiant(e) 2e ou 3e cycle)**

(G*) (POS-63) Optimization and Characterization of Bi-Detector Coincidence Beta-Ray Spectrometry System

Tuesday, 7 June 2022 17:36 (2 minutes)

We present optimization and characterization of a Si-plastic scintillator coincidence beta-ray spectrometer. Recent recommendation to lower the dose limit for the lens of the eye by International Commission on Radiological Protection posed new health physics requirement in the country. Beta-ray dosimetry is of great importance for nuclear industries, particularly during the maintenance periods. The beta-ray spectral data is most fundamental and vital information for accurate beta-ray dosimetry for mixed beta-gamma fields that are often encountered during the nuclear maintenance work. To this end, a Si-Plastic scintillator coincidence beta-ray spectrometer has been developed. The spectrometer can collect pure beta-ray spectra by rejecting the gamma-ray detection events through coincidence. The pulse height and arrival time of each detector signal was processed by a compact digital system and was collected in list mode. A recent upgrade in the digital processor enabled the spectrometer to cover the entire beta energy range of interest. The responses of the spectrometer to beta and gamma were characterized by experiments and Monte Carlo simulations. Spectral measurements under beta-gamma mixed fields with various beta and gamma count rates using ^{90}Sr and ^{137}Cs sources were executed as the evaluation of the system performance. The coincidence beta spectrum was quite stable and consistent in most energy region with the increase of the gamma count rate for a fixed beta field. Development of a real-time spectrum analysis method is currently underway.

Primary author: Mr SUN, Ruoyu (McMaster University)

Co-authors: Dr ATANACKOVIC, Jovica (Ontario Power Generation); Dr HANU, Andrei (Bruce Power); Prof. BYUN, Soo Hyun (McMaster University)

Presenter: Mr SUN, Ruoyu (McMaster University)

Session Classification: DNP Poster Session & Student Poster Competition (4) | Session d'affiches DPN et concours d'affiches étudiantes (4)

Track Classification: Technical Sessions / Sessions techniques: Nuclear Physics / Physique nucléaire (DNP-DPN)