



Canadian Association
of Physicists

Association canadienne
des physiciens et physiciennes

Contribution ID: 3511 Type: **Poster Competition (Graduate Student) / Compétition affiches (Étudiant(e) 2e ou 3e cycle)**

(G*) (POS-65) A Monte Carlo simulation of the feasibility of detecting bone tungsten using X-ray fluorescence

Tuesday, 7 June 2022 17:36 (2 minutes)

Title: A Monte Carlo simulation of the feasibility of detecting bone tungsten using X-ray fluorescence

Authors: Sajed Mcheik(1,*) and Ana Pejovic-Milic(1,)

Affiliations: Department of Physics, Ryerson University, Toronto, ON, Canada, M5B 2K3;

E-mail address of the corresponding author: smcheik@ryerson.ca

An increased number of studies are introducing use of tungsten in medicine in the form of sodium tungstate as an antidiabetic medicine [1], and tungsten nanoparticles as a contrast agent for CT scanning [2] or enhancers of cancer therapy [3]. On the other side, human exposure to tungsten could lead to adverse health effects, including tumour promotion, pulmonary dysfunction, or immune dysfunction [4]. Therefore, it is timely to develop a diagnostic tool to monitor medical exposure to tungsten.

To address this need, we propose developing a robust non-invasive technique to detect bone tungsten in vivo based on the x-ray fluorescence (XRF). A HPGe detector along with homogenous bone phantoms were modeled using Monte Carlo software TOPAS, 3.3 version. A cylindrical shape bone phantom constituted of percent mass as $\text{Ca}_{10}(\text{PO}_4)_6(\text{OH})_2$ [5] (2.7 cm diameter and 8 cm height) were modeled simulating human tibia measurements. The simulation model generated XRF spectra using 109 particles, which were then analyzed to decide on the best excitation source and geometry to optimize the detection limit

The TOPAS simulation showed that Cd-109 is a potential excitation source to detect tungsten in tibia with a minimum detection limit equal to 0.3 ppm W/Ca for 180-degree geometry.

Reference:

[1] Bertinat, R., Westermeier, F., Gatica, R., & Nualart, F. (2019;2018;). Sodium tungstate: Is it a safe option for a chronic disease setting, such as diabetes? *Journal of Cellular Physiology*, 234(1), 51-60.

[2] Jakhmola, A., Anton, N., Anton, H., Messaddeq, N., Hallouard, F., Klymchenko, A., Mely, Y., & Vandamme, T. F. (2013;2014;). Poly- ϵ -caprolactone tungsten oxide nanoparticles as a contrast agent for X-ray computed tomography. *Biomaterials*, 35(9), 2981-2986

[3] Wang, R., Cao, Z., Wei, L., Bai, L., Wang, H., Zhou, S., Ma, Q. (2020). Barium tungstate nanoparticles to enhance radiation therapy against cancer. *Nanomedicine*, 28, 102230-102230.

[4] Bolt, A. M., & Mann, K. K. (2016). Tungsten: An emerging toxicant, alone or in combination. *Current Environmental Health Reports*, 3(4), 405-415.

[5] Da Silva E, Kirkham B, Heyd D V and Pejović-Milić A.(2013) Pure hydroxyapatite phantoms for the calibration of in vivo x-ray fluorescence systems of bone lead and strontium quantification *Anal. Chem.* 85 9189-95

Primary author: Mr MCHEIK, Sajed (Ryerson University)

Co-author: Dr PEJOVIC-MILIC, Ana (Ryerson University)

Presenter: Mr MCHEIK, Sajed (Ryerson University)

Session Classification: DAPI Poster Session & Student Poster Competition (4) | Session d'affiches DPIA et concours d'affiches étudiantes (4)

Track Classification: Technical Sessions / Sessions techniques: Applied Physics and Instrumentation / Physique appliquée et de l'instrumentation (DAPI / DPAI)