



iWoRiD 2022

23rd International Workshop on Radiation Imaging Detectors

26 – 30 June 2022

Riva del Garda, Italy

Contribution ID: 156

Type: **Poster**

Design and preliminary performance of scintillators-based unmanned aerial vehicle for low-cost remote radiation detection

Wednesday 29 June 2022 17:49 (1 minute)

Unmanned aerial vehicles (UAVs) provide an efficient method of remotely sensing environments that humans cannot approach with conventional aircraft due to serious hazard or access limit. The use of UAVs has been suggested as suitable solution in numerous disciplines, including wildfire thermal imaging, radiological survey and radiation activity monitoring since the Fukushima nuclear accident. Many researches have adopted different types of UAVs and radiation detection sensors and have proposed various radiation monitoring systems and data management systems in a site.

In this work, we propose a novel method of using a commercially available drone for aerial radiation detection, which consists of a high-resolution image camera for environmental observation. We have designed and developed the low-cost and remote radiation monitoring system with different types of scintillating screens for radiation detection tasks. A latest compact drone (model: mavic 2 pro) with 322(L) x242(W) x84(H) mm dimension consists of CMOS image array for vision imaging acquisition. Different commercial scintillation screens such as CsI:Tl and powder Gd₂O₂S:Tb materials were applied to measure the X-ray exposure dose and gamma ray activity. The various design parameters such as scintillator types and radiation types were selected and investigated for preliminary possibility under practical X-ray and gamma-ray condition.

For evaluation and optimization of the X-ray image device characterization, different configuration parameters are investigated. The characteristics of the CMOS image sensors with and without scintillator in a drone, such as dose response linearity, dose rate dependence, and minimum detectable activity were evaluated. This result has demonstrated that the unmanned aerial vehicle with a camera and scintillator can be used as a low sensitivity dose rate meter.

Figure. 1. Design concept of a scintillator-based unmanned aerial vehicle (left) and a histogram distribution of an X-ray image (right).

Authors: Mr BAK, Geunho; Mrs KIM, Geonhee; Mrs PARK, Jung Hyun; Prof. SEO, Chang-Woo; Mr SONG, Hyunchul; Mr KANG, Juwan; Mr KIM, Minsung; Mrs KIM, Nahyun; Mr LEE, Hunwoo; Mrs LIU, Wenxuan; Mrs SIM, Jiyong

Co-authors: Prof. JO, Hyosung; Mr CHO, Hyosung; CHA, Bo Kyung

Presenter: Mrs SIM, Jiyong

Session Classification: Poster