



Contribution ID: 249

Type: **Talk (preferred)**

Fluorescence-based Fibre Optic Sensor for Hydrogen Sulphide Detection

Friday 16 December 2022 09:15 (15 minutes)

A fluorescence-based fibre optic sensor has been developed to detect hydrogen sulfide. Optical fibre functionalized with HS-sensitive fluorophore shows an increase in the fluorescence emission upon reaction with HS, the similar behaviour to when fluorophore is dissolved in the solution.

Primary author: BAGHAPOUR, Shaghayegh (Laser Physics and Photonic Devices Laboratories, University of South Australia, SA 5095, Australia)

Co-authors: Prof. PLUSH, Sally (Clinical and Health Sciences, University of South Australia, SA 5000, Australia); Prof. AFSHAR VAHID, Shahraam (Laser Physics and Photonic Devices Laboratories, University of South Australia, SA 5095, Australia); Dr WARREN-SMITH, Stephen (Future Industries Institute, University of South Australia, SA 5095, Australia); Dr ZHANG, Wen Qi (Laser Physics and Photonic Devices Laboratories, University of South Australia, SA 5095, Australia)

Presenter: BAGHAPOUR, Shaghayegh (Laser Physics and Photonic Devices Laboratories, University of South Australia, SA 5095, Australia)

Session Classification: Joint session: AIP-BMP / COMMAD

Track Classification: AIP Congress: AIP: Biomedical and Medical Physics