CERN Nº EDMS REV. VALIDITE 2994766 0.3 **U.** Approval CH-1211 Genève 23 Suisse RÉFÉRENCE ATS-DO SAFETY CERN

> CHNOLOGY DEPARTMENT

> > Date: 8. 11. 2023



Dès qu'imprimé, la validité de ce document est non contrôlée. Vérifier sa validité sur l'EDMS avant de l'utiliser.



 REV.
 VALIDITE

 0.3
 U. Approval

Page 2 de 9

	S	SUIVI D	ES MODIFICATIONS
Nº DE RÉV.	DATE	PAGES	DESCRIPTIONS DES MODIFICATIONS
0.1	8.11.23	All	First draft after online discussion on 24. 10. 23 and information collection
0.2	10.11.23	All	Version for engineering check (CERN organisers and instructors)
0.3	15. 11. 23	All	Comments included, version for approval

## Contents

1. Inti	roduction3
2. Exp	eriments: Location, Organisation, Equipment3
2.1 H	igh-Temperature Superconductors
2.1.1	Location
2.1.2	Organisation3
2.1.3	Equipment4
2.2 M	agnetic Measurements
2.2.1	Location4
2.2.2	Organisation5
2.2.3	Equipment5
3. Наг	ard Register and Simple Mitigation Methods
4. <b>K</b> ISI	RASSESSMENTS for Particular Hazaras
<b>4.1 0</b>	xygen Deficiency Hazard and Mitigation7
4.1.1	Screening estimate7
4.1.2	Time-dependent Estimate (more accurate)7
4.1.3	Mitigation Measures7
4.2 El	ectrical Risk Assessment and Mitigation
4.3 M	agnetic Risk Assessment and Mitigation
4.3.1	Electromagnets
4.3.2	Permanent Magnets8
5. Ref	erences





### Page 3 de 9

# 1. Introduction

The CERN Accelerator School (CAS) organises a specialised course on Course on Normal- and Superconducting Magnets in St. Pölten, Austria 19. 11. to 2. 12. 2023. (<u>https://indico.cern.ch/event/1227234/</u>)

During the two-week course with 96 students and up to 24 instructors , hands-on sessions with demonstration experiments on high-temperature superconductors (HTS) and magnetic measurements will be held in two specially prepared rooms. The cryogenic fluid employed for the superconducting magnets is liquid nitrogen ( $IN_2$ ).

This document lists the electrical, magnetic, cryogenic and oxygen deficiency hazards during the experiments and makes recommendations for mitigation.

# 2. Experiments: Location, Organisation, Equipment

The CAS takes place in Cityhotel D&C in St. Pölten.

## 2.1 High-Temperature Superconductors

### 2.1.1 Location

The "Waldviertel" conference room is reserved for the cryogenic experiments on HTS. The approximate size of the room is 12 m \* 8.5 m \* 4 m (Volume = 408 m<sup>3</sup>), one of the long sides has windows with an estimated surface of 20 m<sup>2</sup>.

At the beginning of the school, liaise with the Hotel reception to clarify evacuation paths from the meeting rooms and assembly points, and obtain an emergency telephone number (ambulance, fire brigade).

Obtain a small first-aid kit from the Hotel reception or a local pharmacy and add an ointment for burns (it will alos calm cold burns from IN2) and have it at hand in the laboratory.

## 2.1.2 Organisation

Five different experiments are set-up on one table each (see Annex 1). At each table, a group of five students performs the experiments during approximately one hour, then they change place with a different group. At least on instructor per table is always available to assist. It is assumed that all five experiments are active at the same time.





## 2.1.3 Equipment

### 2.1.3.1 Cryogenic Equipment

Small transport dewars with a capacity between 5 L and 10L will be used to transport  $IN_2$  from the transport dewars stored in free air to the experiment tables. The IN2 will be poured into simple cryostats, open to air, by instructors. It is assumed that max. 50 L of  $IN_2$  will be present in the room at the same time.

Table	Dewar Sizes	Quantity used	Responsible
1	4 * 2.5 L		CERN
2	1 L, 2.5 L		U Southampton
3	8 L, 4 L	< 10 L / day	CIEMAT
4	10 L	20 L / day	U Wien
5	Combination of the	experiment from table 1-4	·

### 2.1.3.2 Electrical Equipment

All electrical equipment is CE marked. Voltages are restricted to the Very Low Voltage range (Fr: TBT; U < 50 V)

Device	Make	Max range	Responsible
Current source	Keithley 2460	<i>I</i> < 100 mA, <i>U</i> < 100 V	CERN
		Voltage locked to $< 50 V$	
Safety extra-low voltage	Delta Electronica	<i>I</i> < 400 mA, <i>U</i> < 15 V	CERN
(SELV) power supply	SM15-400		
Power Supply	Agilent E3632A	<i>I</i> < 4 A, <i>U</i> < 30 V	U Southampton
Data Logger	Agilent 34970A	N. A.	U Southampton
Precision current source	Lakeshore 120	<i>I</i> < 100 mA, <i>U</i> < 14 V	U Wien

## 2.2 Magnetic Measurements

### 2.2.1 Location

A room is reserved in a Technical School in the vicinity of the Hotel for the experiment of magnets.

At the beginning of the Accelerator School classes, liaise with the Technical School management to clarify evacuation paths from the meeting rooms and assembly points, and obtain an emergency telephone number (ambulance, fire brigade)

Obtain a small first-aid kit from the Technical School management or a local pharmacy and have it at hand in the laboratory.



#### Page 5 de 9

## 2.2.2 Organisation

Five tables with experiments are prepared, housing three different experimental scenarios. At each table, a group of five students performs the experiments during approximately one hour, then they change place with a different group. At least on instructor per table is always available to assist. It is assumed that all five experiments are active at the same time.

## 2.2.3 Equipment

### 2.2.3.1 Magnetic Equipment

Table	Magnet	Electrical characteristics	Magnetic Characteristics
1-3	Air-cooled quadrupole	$I_{max} = 10 \text{ A}$	$L = 32 \text{ mH}, E_{mag} = 1.6 \text{ J}$
	PXMQNAHNAP	R = 340 mΩ	Max. pole tip field 73 mT
4	Air-cooled dipole	$I_{max} = 6 \text{ A}$	$L = 4 \text{ mH}, E_{mag} = 0.07 \text{ J}$
	PXMCCATWAP		Max pole tip field 6 mT
5	1 Halbach-type Linac 4 quadrupole with permanent magnets	N.A.	Pole peak field 230 mT
	Permanent magnets: NdFeB and SmCo	N. A.	Pole tip field > 100 mT.

#### 2.2.3.2 Electrical Equipment

All electrical equipment is CE marked. Operational voltages are restricted to the Very Low Voltage range (Fr: TBT; U < 50 V). All equipment is supplied by and under responsibility of CERN.

Device	Make	Max range
3 Desktop Power supplies	TTi TSX1820P	I < 10 A, U < 20 V
Power Amplifier	Kepko 72 V 6 A	I < 6 A, U < 76 V
Power Amplifier	Kepko 36 V 6 A	I < 6 A, U < 36 V



REV. VALIDITÉ 0.3 U. Approval

Page 6 de 9

# **3. Hazard Register and Simple Mitigation Methods**

Hazar	d	Mitigation	Person responsible of the measure
Cryogenic Fluid: IN <sub>2</sub> (T = 77 K)		During the handling of IN <sub>2</sub> , all participants wear suitable low-temperature gloves and safety goggles.	Amalia Ballarino
Oxygen Deficiency Hazard (ODH)		Detailed assessment in section 4.1	Amalia Ballarino
Transport and storage of $IN_2$		Two storage dewars (capacity 250 L) are stored outside but protected from rain. Instructors fill small transport dewars (capacity between 1 L and 10 L) from the storage dewars and carry them to the experimental room.	Amalia Ballarino
Electricity	A	Detailed assessment in section 4.2	Marco Buzio
		Detailed Assessment in Section 4.3	Marco Buzio
Magnetic Fields		Both experimental rooms are inaccessible to wearers of implanted medical devices. Place the corresponding sign on the doors.	Marco Buzio
Noxious chemical substances		<b>Soldering media:</b> Only instructors are permitted to solder, using a soldering station with local exhaust ventilation.	Marco Buzio
Trip and Fall	A	All electrical cables on the floor shall be secured with black-yellow tape	Amalia Ballarino, Marco Buzio
Mechanical Hazards		All parts, components and samples foreseen to be manipulated by CAS students shall be free from sharp corners The low levels of torque and speed of the rotating coil system are not deemed to be a significant hazard. During operation, students shall be instructed to keep a min. distance of 1 m.	Marco Buzio



#### Page 7 de 9

## 4. Risk Assessments for Particular Hazards

## 4.1 Oxygen Deficiency Hazard and Mitigation

The following estimates show that the average oxygen concentration in the room will not fall below 19 %.

The room "Waldviertel" will be reserved for experiments on HTS, using the cryogenic fluid liquid nitrogen (IN<sub>2</sub>)., The room's volume  $V = 408 \text{ m}^3$ . For all following estimates the general assumption is instantaneous mixing of evaporated gas with the air in the toom.

### 4.1.1 Screening estimate

For a conservative screening estimate, it is assumed that every table uses 10 L of  $IN_2$  per experiment. At the beginning of each phase, **50 L IN\_2** are present in the room. If they would evaporate at once, the resulting  $N_2$  gas at room temperature would occupy 35 m<sup>3</sup>, by displacing air. The remaining concentration of  $O_2$ ,  $c(O_2)$  is [1]:

$$c(0_2) = 21\% \left(1 - \frac{35}{408}\right) = 19.2\%$$

This value is larger than 18 % and therefore acceptable.

### 4.1.2 Time-dependent Estimate (more accurate)

In reality,  $IN_2$  will evaporate at a variable rate over the duration of one hour. The critical value of  $c(O_2, t)$  is therefore time deponent. A more realistic estimate includes the evaporation rate of  $IN_2$ , and the natural air exchange in the room. A conservative approach is that all  $IN_2$  will evaporate during the first 15 minutes when the superconductors are cooled down. The evaporation rate of  $IN_2$  is 3.33 L/min, leading to an evaporated gas flow of  $R = 2.33 \text{ m}^3/\text{min}$ . In a typical room, the air exchange rate is twice per hour, in the  $V = 408 \text{ m}^3$ - room above,  $Q = 13.6 \text{ m}^3/\text{min}$ . The time-dependent concentration of  $O_2$  is [1]:

$$c(0_2, t) = \frac{21\%}{Q+R} \left\{ Q + R \cdot \exp\left(-\frac{Q+R}{V}t\right) \right\}$$

Entering the values for V, Q and R one obtains at t = 15 minutes a value of  $c(O_2) = 19.6\%$ . This value is also in the safe range of oxygen concentration.

#### 4.1.3 Mitigation Measures

The conclusion of the previous estimates is that there is no intrinsic oxygen deficiency hazard during the proposed experiments. The following mitigation measures are nevertheless mandatory to:

- Monitor local concentration of O<sub>2</sub> with portable ODH detector (three at representative locations in the room, for example on some of the working tables).
- Make a full air exchange every hour, when students change table, by opening all windows and the door of the room widely for 5 minutes.



## 4.2 Electrical Risk Assessment and Mitigation

All electrical devices (transformers, power converters, multimeters etc.) bear a CE-mark. The experiments use voltages U < 50 V, i.e. in the **very low voltage range** where the skin resistance protects from further harm.

Only instructors are allowed to connect or disconnect normal-conducting magnets from the power supplies, (see below under 4.3.1.)

To prevent an accidental disconnection of cables, they will be either of type Mammut screw connection, or they will be secured with a cable tie.

If metallic conductors are exposed, they shall be protected by Kapton tape.

### 4.3 Magnetic Risk Assessment and Mitigation

Magnets in both laboratories (HTS and Superconductivity, Magnetic Measurements) cause flux densities more than 0.5 mT. **Persons with implanted medical devices (e.g. pacemaker or insulin pump) must not enter the laboratories**. The standard warning sign will be posted on the doors to the laboratory rooms.

#### 4.3.1 Electromagnets

The electromagnets store a small amount of electromagnetic energy (see table under 2.2.3.1.). This energy is released in case of an abrupt connection or disconnection of the magnet, and therefore, on the electrical terminals an induced voltage will appear. The energy and voltage are, however, not able to create an electrical arc in air.

The connection and disconnection of electromagnets shall only be performed.

- By instructors
- When the power supply of the respective magnet is in position zero or off.

### 4.3.2 Permanent Magnets

Projectile effect: the strong permanent magnets included in the setup may generate a peak field up to 400 mT generate and magnetic forces up to 100 kg ("aimant de la mort" [3]) and therefore do represent a risk of projectile effect, as well as pinching body parts and/or splintering when manipulated. Mitigating measures: a) the magnets shall be kept normally enclosed in a box such as that no significant magnetic force can be generated outside; b) the magnets generating more than 15 kg of force shall be operated only by trained CERN staff; c) the students shall receive appropriate training to operate smaller magnets, including the obligation to keep a minimum 0.5 m distance from other magnets, any steel part or electronic equipment, or potentially sensitive personal items (credit cards, mechanical watches etc)

## 5. References

[1] Thomas Otto, Safety for Particle Accelerators, Section 2.3. Springer Nature Switzerland, Cham 2021

[2] CERN General Safety Instruction: Non-Ionising Radiation, GSI-NIR-1

[3] Source: <u>Supermagnete.fr</u>

Annex 1: Summary of High-Temperature Superconductor Experiments

*				
	Table 1 (Prepared by CERN)	Table 2 (Prepared by SOTON)	Table 3 (Prepared by CIEMAT)	Table 4 (Prepared by Un. of
	Properties of Superconductors			Vienna)
Magnets and	- REBCO Tape - PEBCO Builk	- REBCO Tape - No magnets	- Laptop Computer - Software Ouick Field Student version	- LN2 containers: 1 container of 10 l. To he refilled regularly At least 20 l are
cryostatic	- Dermanent magnet	- I N7 containers 1 container 2 5 I and 1	- 0 +HTS mile (manufactured 2)	perfect each day
equipment	- Permanent magnet in U-chaned	container 1.0.1	provided by CIFMAT)	- REBCO bulk
	configuration	- Glass prvostat with vacuum insulation	- 3* Polystyrene Boxes (Cryostat	- Stvrofoam box
	- LN2 4 containers 2.5 I each. total of 10	(from CERN)	provided by CIEMAT)	- HTS coil
	[]	- Liquid nitrogen open cryostat (1 units	- 2* LN2 Dewar for filling the cryostat (8	<ul> <li>Various permanent magnets</li> </ul>
	- Glass cryostat with vacuum insulation	from CERN)	& 4 litter, respectively)	- Varnish insulated copper wire
	<ul> <li>Liquid nitrogen open cryostat (2 units)</li> </ul>		<ul> <li>Demonstration magnetic bearing kit</li> </ul>	-Teflon tape
			(manufacturing & provided by CIEMAT)	
Electrical	- Safety extra-low voltage (SELV) power	- Agilent E3632A Power Supply, 30 V, 4A	- DC power supply (8V- 125 A)	- Two miliampere current sources
Instrumentation	supply (2 units): Delta Electronica	<ul> <li>Agilent 34970A Data Logger, no power</li> </ul>	Needed: 1 socket 10A single-phase	(Lakeshore 120 precision current
	SM15-400, 15 V, 400 A	output	- Millivoltmeter	sources: max 100 mA, max 14 V)
	- Keithley 2460, 100 V, 1.05 A, the	<ul> <li>A laptop computer</li> </ul>	<ul> <li>DC Amperemeter up to 125A</li> </ul>	- Lock-In-Amplifier
	voltage will be limited to below 48 V	Needed: 3 outlets 10 A monophase	<ul> <li>Thermometer (77K-300K)</li> </ul>	- Voltmeter
	and front panel locked	- Datalogger (1 unit)	- Data logger	- Hall probe
	Needed: 2 outlets 16 A three-phase	<ul> <li>Power supply (1 units)</li> </ul>		- Temperature sensor
	<ul> <li>Cryogenic Pt sensors</li> </ul>	- laptop (1 unit)	<ul> <li>Laptop with Labview</li> </ul>	<ul> <li>Computer with USB ADC converter</li> </ul>
	- Hall probe		- Screen	
	<ul> <li>Four minicomputers for DAQ</li> </ul>			-Heating cartridge
	Needed: 2 outlets 10 A monophase		Needed: up to 3 sockets 5A single-	
	- Multimeters with shunt (2 units)		phase	
	- Nanovoltmeter (4 units)			
	- Source measure unit (2 units)			
	- Lakeshore 224 (2 units			
Exercises	-critical current of REBCO tape at 77 K	- Resistive transition as a function of	- The process of designing a HTS	<ul> <li>Inductive measurement of</li> </ul>
and	and in self-field	temperature of REBCO tape and	magnet (pptx presentation)	superconducting transition temperature
Montenation	<ul> <li>anisotropy of critical current of</li> </ul>	Bi2223 tape	- HTS Coil fabrication (video	<ul> <li>Flux transformer</li> </ul>
Medsurements	REBCO tape: use of U-shaped magnet		presentation)	<ul> <li>Persistent mode operation, influence</li> </ul>
(Summary)	at 77 K		- Cooling down process of the HTS coil	of joint resistance
	- critical temperature of REBCO tape		controlled by thermometer	<ul> <li>Flux trapping in magnet</li> </ul>
	<ul> <li>Resistive transition of REBCO tape</li> </ul>		- V-I curve measuring of the HTS coil at	<ul> <li>Flux pumping</li> </ul>
	and quench propagation		77 K using voltage taps placed in the	
	<ul> <li>Field-cooled and zero field-cooled</li> </ul>		coils	
	experiment with REBCO bulk and		<ul> <li>Optional: Operating a magnetic</li> </ul>	
	permanent magnet		bearing demonstrator based on a Field-	
	<ul> <li>electrical resistivity of selected</li> </ul>		cooled process*	

2994766

0.3





Note: On a 5<sup>th</sup> table, some of the experiments of tables 1 – 4 will be repeatedly installed.

materials (copper, stainless steel and REBCO tape) at room temperature and

at 77 K

VALIDITĖ