

# **International workshop "Positronium - from Quantum Physics to Medical Applications"**

Wednesday, 26 April 2023 - Friday, 28 April 2023

Split, Croatia

## **Book of Abstracts**



# Contents

Welcome and registration . . . . .	1
Opportunities to do experiments in an underground laboratory . . . . .	1
Positronium imaging with modular J-PET for medical diagnostics and basic research in physics . . . . .	1
Nuclear medical imaging based on entangled photon pair . . . . .	1
Quest for High Energetic Entanglement from Positronium Decays: A Rigorous Study . .	2
Nanoparticles for theranostics with PET and MRI . . . . .	2
Development of Compton-PET hybrid camera for multi-tracer imaging in nuclear medicine . . . . .	2
Single Layer Gamma-Ray Polarimeter for Medical Imaging Applications and Fundamental Physics Research . . . . .	3
Investigation of decoherence of annihilation quanta . . . . .	3
Investigation of a Single-plane Compton gamma camera as a radiation imager . . . . .	3
Development and characterization of the PET demonstrator with measurement of polarization correlations . . . . .	4
Search for the CP symmetry violation in the OPSVIO project . . . . .	4
Measurements of the polarization of photons in ortho-positronium annihilation . . . . .	5
An overview of basic concepts and formulas in simulations of the Compton scattering . .	5
Recent advances of positron emission tomography image reconstruction . . . . .	5
Enhanced Positronium formation and Annihilation Localization with nano-scale magnetization . . . . .	6
Medical physics in everyday practice of nuclear medicine . . . . .	7



1

## Welcome and registration

Positronium in fundamental investigations / 2

## Opportunities to do experiments in an underground laboratory

**Corresponding Author:** catalina.oana.petrascu@cern.ch

I shall discuss opportunities to do underground experiments by presenting our investigations of possible departures from the standard quantum mechanics' predictions at the Gran Sasso underground laboratory in Italy.

In particular, with refined radiation detectors we are searching signals predicted by the dynamical collapse models (spontaneous emission of radiation) which were proposed to solve the "measurement problem" in quantum physics, and signals indicating a possible violation of the Pauli Exclusion Principle.

I shall discuss our recent results which ruled out the natural parameter-free version of the gravity-related collapse model. I shall then present more generic results on testing CSL (Continuous Spontaneous Localization) collapse models and discuss future perspectives.

Finally, I shall briefly present the VIP experiment, with which we look for possible violations of the Pauli Exclusion Principle by searching for "impossible" atomic transitions, and comment the impact of this research in relation to Quantum Gravity models.

I shall take this opportunity to stimulate discussions about possibilities to use positronium in future underground experiments.

Positronium in fundamental investigations / 3

## Positronium imaging with modular J-PET for medical diagnostics and basic research in physics

**Corresponding Author:** p.moskal@uj.edu.pl

The Jagiellonian Positron Emission Tomograph (J-PET) is the first PET scanner based on plastic scintillators.

It is designed to measure momentum vectors and the polarization of photons originating from the decays of positronium.

In combination with the newly invented positronium imaging method, J-PET enables the study of discrete symmetries in positronium without the use of magnetic fields. We will present the latest results of P, T, CP, and CPT symmetry studies with the J-PET detector (Nature Communication 12, 5658 (2021)) as well as explain the method of positronium imaging that enables imaging of positronium properties in living organisms (Science Advances 7, eabh4394 (2021), Nature Reviews Physics 1, 527 (2019)). In the talk, the method of positronium imaging in living organisms and the first in-vivo positronium images of humans obtained with the J-PET tomograph will be also presented and explained.

Positronium in medical imaging / 4

## **Nuclear medical imaging based on entangled photon pair**

**Corresponding Author:** shimazoe@g.ecc.u-tokyo.ac.jp

Research activities related to novel nuclear medicine instrumentation in the university of Tokyo will be introduced. Especially Compton and PET hybrid scanner, quantum entanglement PET and simultaneous imaging and sensing based on cascade entangled photons will be discussed together with its readout electronics and image reconstruction.

**Positronium in fundamental investigations / 5**

## **Quest for High Energetic Entanglement from Positronium Decays: A Rigorous Study**

**Corresponding Author:** beatrix.hiesmayr@univie.ac.at

Recently, theoretical and experimental papers have been published that put forward different interpretations about the physics of the photons produced in positronium decay. Most of the publications claim that their data prove quantum entanglement. This talk will give an overview of the underlying assumptions and the prospects of the upcoming experiments.

**Positronium in medical imaging / 6**

## **Nanoparticles for theranostics with PET and MRI**

**Corresponding Author:** zdenka.kuncic@sydney.edu.au

In nuclear medicine, cancer theranostics refers to the strategy of combining diagnostic imaging with targeted therapy. In this talk, I will discuss how theranostic efficacy may be enhanced by leveraging the unique properties of nanoparticles. Due to their nanoscale size, nanoparticles can penetrate target tissues and tumour cells, and enhance physico-chemical reaction rates. Additionally, nanoscale geometric confinement confers unique physical properties. Superparamagnetic nanoparticles, in particular, offer a unique platform for radiolabelling with PET isotopes, while also enhancing contrast in Magnetic Resonance Imaging (MRI). Additionally, such nanoparticles, when radiolabelled with a therapeutic isotope, can enhance localised damage to tumour cells, thereby improving therapeutic outcomes.

**Positronium in medical imaging / 7**

## **Development of Compton-PET hybrid camera for multi-tracer imaging in nuclear medicine**

**Corresponding Author:** uenomachi.mizuki.6a@kyoto-u.ac.jp

Simultaneous imaging of PET (positron emission tomography) and SPECT (single photon emission computed tomography) nuclides is difficult in commercial nuclear medicine imaging systems due to their different principles, such as the presence or absence of mechanical collimators. We have proposed Compton-PET hybrid imaging system, which can perform simultaneous PET and SPECT

nuclides imaging by the conventional PET and Compton imaging, without any mechanical collimators. We have developed some prototype cameras using different detectors, such as GAGG-SiPM (silicon photomultiplier) detectors, CeBr3-SiPM detectors and silicon detectors. In the presentation, we will introduce the detail of our imaging system and demonstration results.

## Positronium in fundamental investigations / 8

### Single Layer Gamma-Ray Polarimeter for Medical Imaging Applications and Fundamental Physics Research

**Authors:** Mihael Makek<sup>1</sup>; Tomislav Bokulic<sup>1</sup>; Damir Bosnar<sup>1</sup>; Darko Grošev<sup>2</sup>; Ivica Frišić<sup>1</sup>; Ana Marija Kožuljević<sup>3</sup>; Zdenka Kuncic<sup>4</sup>; Luka Pavelić<sup>5</sup>; Petar Zugec<sup>1</sup>; Siddharth Parashari<sup>1</sup>

<sup>1</sup> *Department of Physics, Faculty of Science, University of Zagreb*

<sup>2</sup> *University Hospital Centre Zagreb*

<sup>3</sup> *Department of Physics, Faculty of Science, University of Zagreb*

<sup>4</sup> *University of Sydney, Australia*

<sup>5</sup> *Institute for Medical Research and Occupational Health*

**Corresponding Author:** makek@phy.hr

We will present an overview of the activities undertaken with the experimental system based on single layer gamma-ray polarimeter. This modular system consist of 16 position sensitive scintillator matrices read out by silicon photomultipliers. We have shown that these simple detectors can successfully measure the polarization of gamma rays via internal Compton scattering. Owing to its modularity the system can be exploited in various setups in fundamental research and medical applications. We will present three such setups: the first one for the measurements in Positron Emission Tomography, where the benefit of using the gamma-ray polarization was investigated, the second where the implications of decoherence of annihilation quanta were explored and the third, where triple-coincident measurement of gamma rays from ortho-positronium decay were done.

## Positronium in fundamental investigations / 9

### Investigation of decoherence of annihilation quanta

**Authors:** Siddharth Parashari<sup>1</sup>; Damir Bosnar<sup>1</sup>; Zdenka Kuncic<sup>2</sup>; Mihael Makek<sup>1</sup>

<sup>1</sup> *Department of Physics, Faculty of Science, University of Zagreb*

<sup>2</sup> *University of Sydney, Australia*

**Corresponding Author:** siddharth@phy.hr

Recently, different groups have performed measurement of polarization correlations of annihilation quanta after inducing decoherence of one of the gammas by Compton scattering yielding somewhat contradictory results. Watts et al. [Nat. Commun., 12, 2646, (2021)] reported the result hinting at loss of correlation, while Abdurashitov et. al. [Jour. Inst. 17, P03010, (2022)] reported strong correlation at least at small scattering angles. We have used the setup based on Single Layer Gamma Ray Polarimeters to measure the correlation of annihilation quanta after an intermediate Compton scattering under angles ranging from 0 to 50 degree, thus significantly extending the range compared to previous measurements. The results of these investigations will be presented.

## Positronium in medical imaging / 10

## Investigation of a Single-plane Compton gamma camera as a radiation imager

**Authors:** Om prakash Dash<sup>1</sup>; Tomislav Bokulić<sup>2</sup>; Damir Bosnar<sup>3</sup>; Mihael Makek<sup>3</sup>

<sup>1</sup> *Department of Physics, Faculty of Science, University of Zagreb*

<sup>2</sup> *Department of physics, Faculty of Science, University of Zagreb*

<sup>3</sup> *Department of Physics, Faculty of Science, University of Zagreb*

**Corresponding Author:** ompd@phy.hr

Radiation imaging and detection is an outstanding topic in various areas from astroparticle physics, over medical imaging to radiation security. On of suitable detectors which has been researched lately is the Compton camera offering potential advantages such as a wide field of view, the ability to reconstruct 3D images, and with a portable lightweight design due to absence of heavy collimation. We designed and constructed a novel, compact Compton gamma camera whose detector element consists of two scintillator crystals optically coupled by a light guide between them. We used GAGG:Ce scintillators of 3 x 3 x 3 mm<sup>3</sup> and 3 x 3 x 20 mm<sup>3</sup> plexiglass lightguides. Detector elements were placed in an 8 x 8 matrix with a 3.2 mm pitch, separated by ESR reflector. In this configuration the front scintillator layer is acting as the scatterer and the back scintillator layer is acting as the absorber of the Compton scattered gamma radiation, while both are read out by the same silicon photomultipliers (SiPM) array coupled to the back side of the matrix, thus forming a compact single-plane detector. We will report the results of the performed laboratory characterization of the detector.

**Positronium in medical imaging / 11**

## Development and characterization of the PET demonstrator with measurement of polarization correlations

**Authors:** Ana Marija Kožuljević<sup>1</sup>; Siddharth Parashari<sup>2</sup>; Tomislav Bokulic<sup>3</sup>; Damir Bosnar<sup>2</sup>; Darko Grošev<sup>4</sup>; Zdenka Kuncic<sup>5</sup>; Luka Pavelić<sup>6</sup>; Petar Žugec<sup>2</sup>; Mihael Makek<sup>2</sup>

<sup>1</sup> *Department of Physics, Faculty of Science, University of Zagreb*

<sup>2</sup> *Department of Physics, Faculty of Science, University of Zagreb*

<sup>3</sup> *Department of physics, Faculty of Science, University of Zagreb*

<sup>4</sup> *University Hospital Centre Zagreb*

<sup>5</sup> *University of Sydney, Australia*

<sup>6</sup> *Institute for Medical Research and Occupational Health*

**Corresponding Author:** amk@phy.hr

**Positronium in fundamental investigations / 12**

## Search for the CP symmetry violation in the OPSVIO project

**Corresponding Author:** ifriscic@phy.hr

One of three Sakharov conditions needed to explain the matter-antimatter asymmetry problem is the existence of CP symmetry breaking. This was indeed observed in the quark sector but with the magnitude which is not large enough to solve the matter-antimatter asymmetry problem. Since the CP violation is allowed by some leptogenesis models, the search for additional CP symmetry breaking was extended to leptonic sector. A recent experiment involving neutrino oscillations reported the indication of CP violation but with no decisive conclusion.



A complementary approach to search for the CP violation in leptonic sector is to use the tensor polarization of ortho-positronium (o-Ps) induced by an external magnetic field and measure the angular distribution of produced gamma rays. More than one-decade old result showed that there is no CP violation in o-Ps decay at precision level of  $10^{-3}$ . In this presentation we will present the experimental set-up of the OPSVIO project with which we plan to improve the precision level for one order of magnitude compared to the state-of-art result.

**Positronium in fundamental investigations / 13**

## **Measurements of the polarization of photons in ortho-positronium annihilation**

**Corresponding Author:** bosnar@phy.hr

An overview of the previous measurements of the polarization of photons in positronium annihilations, with special emphasis on ortho-positronium annihilations, will be given. Possibilities for the measurements of the polarization of all three photons in ortho-positronium annihilations with single-layer Compton scattering detector systems developed at the Department of Physics in Zagreb as well as the first measurements will be presented.

**Positronium in fundamental investigations / 14**

## **An overview of basic concepts and formulas in simulations of the Compton scattering**

**Author:** Petar Žugec<sup>1</sup>

<sup>1</sup> *Department of Physics, Faculty of Science, University of Zagreb*

**Corresponding Author:** pzugec@phy.hr

We give an overview of basic concepts and formulas used in simulating the Compton scattering. Special emphasis is given to a Klein-Nishina expression for a differential scattering cross section for a single photon. In a context of entangled annihilation photons, a scattering of single (uncorrelated) photons is a reference point for gauging the quantum effects due to the entanglement itself. As such, the uncorrelated scattering of multiple photons must be perfectly characterized when analyzing the experimental data pertaining to the entangled photons. In that, an experimental evidence for the photon entanglement and its effect upon subsequent scattering is to be found in the observed deviations from the uncorrelated scattering.

**Positronium in medical imaging / 15**

## **Recent advances of positron emission tomography image reconstruction**

**Corresponding Author:** dora.varnyu@mediso.com

Positron emission tomography (PET) is a widely used imaging modality that enables the non-invasive visualisation of physiological and biochemical processes in living organisms. However, PET images are inherently noisy and suffer from low spatial resolution, which can limit their diagnostic accuracy

and clinical utility. To address these challenges, numerous image reconstruction methods have been proposed and developed over the past years.

In this presentation, we will discuss the recent trends in PET image reconstruction, focusing on techniques designed to improve image quality, allow the reduction of the injected radioactive dose as well as the acquisition time. Specifically, we will cover dynamic and quantitative PET imaging, total-body PET, time-of-flight (TOF) technology, resolution recovery. Finally, we will discuss the challenges and opportunities in using artificial intelligence (AI) and deep learning (DL) for PET.

## Positronium in medical imaging / 16

### Enhanced Positronium formation and Annihilation Localization with nano-scale magnetization

Developments in radiolabeling superparamagnetic iron oxide nanoparticle (SPIONs) have gained increasing attention for cancer theranostic applications<sup>1</sup>. In a previous study, we demonstrated that the FDA approved SPION Feraheme® (FH) can be radiolabeled with a range of therapeutic and diagnostic isotopes: <sup>64,67</sup>Cu, <sup>90</sup>Y, <sup>177</sup>Lu, <sup>89</sup>Sr, <sup>140</sup>Ba, <sup>99</sup>Mo, <sup>212</sup>Pb, <sup>213</sup>Bi, <sup>111</sup>In, <sup>153</sup>Sm, <sup>161</sup>Tb, <sup>156,157</sup>Eu by the chelate-free heat induced radiolabeling technique<sup>2</sup>. In separate studies, we also demonstrated that radiolabeled FH can enhance dose deposition<sup>3</sup> and that <sup>89</sup>Zr-FH is a highly suitable radio-nanoplatfrom for hybrid PET- MR imaging<sup>4</sup>. Here, we investigate for the first time the effect of magnetized radiolabeled FH SPIONs on dose localization and positron range, as well as and ortho-positronium production from <sup>89</sup>Zr-FH.

A series of [Fe] dilutions of <sup>89</sup>Zr-FH samples was prepared in 10 separate phantom vials (including only <sup>89</sup>Zr). The activity in each phantom was kept constant, A<sub>0</sub> = 3.7 kBq. The phantoms were scanned using a simultaneous clinical PET-MR scanner (3T Biograph mMR). The full width at half maximum of the line spread function were calculated for the PET image data to assess the impact of magnetized FH SPIONs on the spatial resolution. The integrated standard value uptake for a circular region of interest for each phantom scan was calculated to quantify the dose localization.

Results demonstrated the magnetized FH SPIONs improved the spatial resolution of the <sup>89</sup>Zr-FH phantom PET images by  $\approx 17 \pm 1.6$  %, localized the dose by  $\approx 40 \pm 0.9$  % and increased the true and random counts by  $\approx 6$  % and 1% respectively, at a clinical [Fe] FH dose level. Both improvements in spatial resolution and dose localization are due to the nano-scale enhanced magnetic field induced by magnetized FH SPIONs and this has been further confirmed by PET-MR image analysis. In a clinical scenario, enhancing dose localization by 40% may improve the tumor control probability by 40%. Furthermore, the increase in true and random counts may be due to the interaction of positrons within the <sup>89</sup>Zr-FH solution resulting in annihilation via formation of ortho-positronium (in the triplet state, <sup>3</sup>S<sub>1</sub>) and emission of three gammas. Thus, this study further suggests radiolabeled SPION can enhance the production of ortho-positronium. Such triple-coincidences may be processed as a set of double coincidence events by the PET scanner. It would be interesting to follow up our study using emerging total body PET scanners, which with their superior sensitivity and 4 $\pi$  geometry are ideal for triple coincidence detection and potentially for ortho-positronium emission tomography.

#### Reference:

1. Assadi, M., Afrasiabi, K., Nabipour, I. & Seyedabadi, M. Nanotechnology and nuclear medicine; research and preclinical applications. *Hell J Nucl Med* 14, 149–159 (2011).
2. Gholami, Y. H., Akam E A, Caravan P, Wilks M. Q, Xiang-Zuo Pan, Maschmeyer R, Kolnick A, El Fakhri G, Normandin Marc D, Josephson L, Kuncic Z, Yuan H. A Chelate-Free Nano-platfrom for Incorporation of Diagnostic and Therapeutic Isotopes. *Int J Nanomedicine*, in press.
3. Gholami, Y. H., Maschmeyer, R. & Kuncic, Z. Radio-enhancement effects by radiolabeled nanoparticles. *Sci Rep* 9, 1–13 (2019).
4. Gholami, Y. H., Yuan H, Wilks M. Q, Maschmeyer R, Normandin Marc D, Josephson L, El Fakhri G, Kuncic Z. A A radio-nano-platfrom for T1/T2 dual mode and PET-MR hybrid modal imaging. *Int J Nanomedicine*, (under review).

**Positronium in medical imaging / 17**

**Medical physics in everyday practice of nuclear medicine**

Short overview will be given of the possible activities of medical physicist (MP) in nuclear medicine (NM). Starting from various aspects of radiation protection issues stemming from the fact that nuclear medicine deals with open radioactive sources (radiopharmaceuticals). Quality control of instruments such as gamma camera and PET scanner are also an important part of the routine job of MP. Few examples will be given of how MP can apply mathematical modeling and sophisticated software to improve both diagnostic and therapeutic procedures in NM, thus helping medical doctors in solving important clinical questions.