

SILICON CARBIDE DIODES FOR ULTRA-HIGH DOSE RATE DOSIMETRY

Tuesday 20 June 2023 14:20 (20 minutes)

The recent development of FLASH radiotherapy has led to the challenge of developing adequate sensors for active dosimetry in Ultra-High Dose Rate (UHDR) beam delivery. Especially in the case of FLASH electron beams the dose delivery can reach up to several Gy even in a single pulse with a few microsecond duration. The accurate dosimetry of this new UHDR radiotherapy modalities represents a key issue for its clinical translation. In this talk I will present the last results on Silicon carbide p-n diode dosimeters designed for the stringent requirements of FLASH radiotherapy that have been fabricated and characterized in an ultra-high dose rate electron beam.

Primary author: Dr PELLEGRINI, Giulio (Centro Nacional de Microelectrónica (IMB-CNM-CSIC) (ES))

Presenter: Dr PELLEGRINI, Giulio (Centro Nacional de Microelectrónica (IMB-CNM-CSIC) (ES))

Session Classification: SiC