

# A Gaseous Compton Camera for Gamma Imaging

Wednesday 3 July 2024 14:35 (1 minute)

A Gaseous Compton Camera for Gamma Imaging

L. F. N. D. Carramate, R. Mendes, A. Correia, F. Lucas, M. E. Loureiro, V. Bonifácio, J. F. C. A. Veloso, and C. D. R. Azevedo

I3N, Physics Department, University of Aveiro, 3810-193 –Aveiro, Portugal

Compton Cameras have been pointed as a possible Anger Camera competitor due to their enhanced sensitivity in imaging radioactive sources, primarily due to the absence of a mechanical collimator. This advantage holds the potential to, not only, improve the quality of nuclear medicine imaging and, simultaneously, to minimize patients' radiation exposure. Their configuration enables the detection of scattered photons and recoil electrons, as well as their energy determination, which allows the construction of a conic surface indicative of the primary photon's interaction site. By intersecting these conic surfaces, it becomes possible to pinpoint the most probable location of interaction. Within the DRIM group, a Compton Camera based on Gaseous Detectors is being developed [1]. The proposed Compton Camera concept has the potential to increase sensitivity when compared to other solutions due to the ability to detect scattered photons across a  $4\pi$  solid angle using only one detector, unlike other designs that require two detectors. In this scope, a simulation study was conducted with a radioactive point source to evaluate the feasibility of the proposed solution. We will present the study results, particularly, the obtained images, that confirm the set-up viability.

[1] C. D. R. Azevedo et al., "A Gaseous Compton Camera using a 2D-sensitive gaseous photomultiplier for Nuclear Medical Imaging," *Nuclear Instruments and Methods in Physics Research, Section A: Accelerators, Spectrometers, Detectors and Associated Equipment*, vol. 732, pp. 551–555, 2013, doi: 10.1016/j.nima.2013.05.116.

Acknowledgements: The costs resulting from the FCT (Fundação para a Ciência e a Tecnologia, I.P.–Portuguese Foundation for Science and Technology) hiring, L.F.N.D. Carramate, were funded by national funds (OE) in the scope of the framework contract 2022.00387.CEECIND (<https://doi.org/10.54499/2022.00387.CEECIND/CP1720/CT0022>). This work is financed by national funds through FCT - Fundação para a Ciência e a Tecnologia, I.P. and the Ministry of Science, Technology and Higher Education under project LA/P/0037/2020, UIDB/50025/2020, UIDP/50025/2020.

**Primary authors:** CORREIA, Afonso (University of Aveiro); AZEVEDO, Carlos (University of Aveiro); LUCAS, Francisco (University of Aveiro); DAS NEVES DIAS CARRAMATE, Lara Filipa (University of Aveiro (PT)); LOUREIRO, Maria Eduarda (University of Aveiro); MENDES, Rui (University of Aveiro); BONIFÁCIO, Vitor (University of Aveiro); VELOSO, joao (university of aveiro)

**Presenter:** DAS NEVES DIAS CARRAMATE, Lara Filipa (University of Aveiro (PT))

**Session Classification:** Poster Session