Advancements in Charge Detection: A High-Resolution, Wide Dynamic Range Prototype for Ion Beam Monitoring *ICHEP 2024 18-24 July*

Prague

CHEP 2024




NFN




1. The research group background.

Our research group is involved in several space experiments for the direct measurement of cosmic-ray, with one notable example being HERD. HERD, a calorimeter designed for measuring protons and nuclei up to energies of 10^{15} eV

Tentative HERD design




Expected proton flux measured by HERD



2. An ion beam monitor for HERD

4. First Results

In 2022, we utilized the initial prototype during the HERD beam test with ions at the SPS. The primary objective was to identify nuclei and assess the nonlinearity of the LYSO scintillator, which serves as the active material in the HERD CALO.



We implemented a simple analysis strategy:

- Convert ADC -> MIP
- Convert MIP -> Z
- Evaluate if ADC signal > threshold.
- Selection: number of diodes above noise threshold > 3.

Charge peaks (fitted with Gaussian distributions) up to Z ~ 60.

In 2023, within the context of the HERD project, nuclei tagging was implemented to evaluate the performance of the large-scale prototype with nuclei.

Simultaneously, in 2023 we conducted our first test outside the HERD collaboration, in support of AMS-02 beam test. The data collected from this test were utilized for independent charge tagging and evaluation of

2022 HERD beam test with ions at SPS, the collaboration needed:


- a beam monitor to quickly check the composition of the beam,
- a reliable and independent charge measurement.
- The detector has been designed with the following features:
- Ease of mounting and dismounting due to HERD's mechanical constraints
- Prompt online analysis for quick results
- Thin construction to minimize the incidence of fragmented nuclei in the detector
- high dynamic range to measure charge from 1 to ~80 (Pb).

Additionally, a significant challenge was the limited time available before the test.


3. Design of the first prototype



Due to the time restriction, well known components have been employed. The prototype were made of 6 PIN diodes [1] VTH2090 which is a large area (9.2 x 9.2 mm2) PD. Measured depletion depth is ~ 220 um .



A very simple mechanical structure was designed




Capacitance measurements for depletion

10,0 20,0 30,0 40,0 50,0 60,0 70,0 80,0 90,0 100,0 110,0 120,0 130,0 140,0 150,0

Feedback

Network

charge reconstruction efficiency.



Charge peaks for A/Z ~ 2.2 SPS ion beam.

The obtained charge resolution is better 0.3 charge unit from *Z* =4 to *Z* = 75 and just slightly increase with Z.


> Good response linearity is also achieved.

Improved analysis strategy:

- Evaluate the number of consistent diodes with the one considered.
- Selection criterion:: select a maximum of 6 consistent diodes.

Optimized diodes bias:

The bias voltage has been raised to 100 V to maximize the depleted region depth.



New project: BeER 5.



To achieve the necessary high dynamic range, the read-out electronics of the calorimeter photodiode system were employed. Specifically, a custom chip HiDRA, version 2, developed by INFN-Trieste (Italy), was utilized.[2]. Single channel block diagram

Cf2 CSA Reset

CSA Res

G2SEL

– Double gain CSA with automaticgain selection circuitry


- High dynamic range (~ 5×10^5)
- Low power consumption ~ 3.5 mW/chan
- Low noise: ENC ~2500e
- 16 input channels
- Self-trigger system

6. References

69

1. O. Adriani et al 2019 JINST 14 P11004

2. O. Adriani et al 2022 JINST 17 P09002



- event-by-event charge tagging system for off-line analysis for ion beam user like ASM-02.
- on-line monitor of high multiplicity beams (e.g. LNF-BTF).
- application in experiments, e.g. nuclei cross section measurement, heavy ion identification in cosmic rays, ...

We are completely open to suggestions about BeER application and future design.

- We are currently in the process of implementing a new design for our detector::
- new sensors VTH2021 (10x10 mm active area, no package)
- Enlarge the active area of the detector by adding more diodes (matrix 3x3)
- 6 layers with 3x3 matrix
- Adjust the mechanics for the new design, add motorized sliders.
- Adjust the electronics to read-out more channels using HiDRA 2 chip.

