



Canadian Association
of Physicists

Association canadienne
des physiciens et physiciennes

Contribution ID: 4326

Type: Oral (Non-Student) / Orale (non-étudiant(e))

2D Slice Selective Head Sized TRASE MRI by Coil Rotation

Monday 27 May 2024 14:30 (15 minutes)

In this work we present the first low-field TRASE technique capable of encoding 2D axial slices without switching gradients of the main magnetic field (B_0). TRASE is an MR imaging technique that utilizes phase gradients within the radiofrequency (RF) fields to achieve k-space encoding. In doing so, TRASE does not require as many technologies of the main magnetic field, significantly reducing the cost and size of the overall system. The TRASE encoding principle ideally requires two and four different RF phase gradient fields for 1D encoding and 2D imaging respectively. Preventing interactions between these RF transmit coils has been the primary challenge, especially for 2D imaging. To address this problem, we constructed a head sized TRASE coil pair capable of 1D encoding any transverse axis. By method of rotation, the encoding axis can be changed, allowing a full 2D k-space acquisition in a radial spoke fashion. This radial TRASE technique requires half the RF transmit coils and accompanying RF electronics than typical cartesian TRASE imaging. As a first demonstration of this technique, a head sized coil pair was constructed and experimentally verified on a uniform 8.65 MHz bi-planar permanent magnet with a constant B_0 gradient used for slice-selection. Decoupling of the two transmit coils is performed geometrically and a parallel-transmit system (PTx) is presented as a method to reduce any residual coupling. This work demonstrates that 2D slice-selective imaging is feasible without the use of any B_0 switched gradients.

Keyword-1

Medical Imaging

Keyword-2

MRI

Keyword-3

Primary authors: Mr SEDLOCK, Christopher (University of Alberta); Dr PURCHASE, Aaron (University of Alberta); Dr TOMANEK, Boguslaw (University of Alberta); Prof. SHARP, Jonathan (University of Alberta)

Presenter: Prof. SHARP, Jonathan (University of Alberta)

Session Classification: (DPMB) M2-3 Low-field Magnetic Resonance Imaging | Imagerie par résonance magnétique à bas champ (DPMB)

Track Classification: Technical Sessions / Sessions techniques: Physics in Medicine and Biology / Physique en médecine et en biologie (DPMB-DPMB)