



Canadian Association  
of Physicists

Association canadienne  
des physiciens et physiciennes

Contribution ID: 4633 Type: **Poster Competition (Graduate Student) / Compétition affiches (Étudiant(e) 2e ou 3e cycle)**

## **(G\*) (POS-47) Navigating Radioguided Cancer Surgery Using Principles of Gamma Photon Interactions**

*Tuesday 28 May 2024 18:35 (2 minutes)*

Breast cancer is the leading cause of cancer in women worldwide and surgery to remove the tumour and stage the cancer is a crucial component of most treatment plans. Radioguided surgery using a hand-held gamma probe that counts gamma photons is a common technique that allows surgeons to locate non-palpable, radiolabeled lesions in the operating room. While gamma probes are effective for detection of low-energy radiotracers, increased scattering at higher energies degrades the probe's resolution and limits the use of high-energy radiotracers, such as positron emitters. Understanding the physics of photon interactions and their influence on the shape of a detected gamma-ray energy spectrum, we hypothesized that a machine learning model could analyze the energy spectrum recorded by a gamma probe and predict the location from which the gamma photons originated. As such, the goal of the study was to assess how well machine learning improves a gamma probe's ability to localize high-energy radiotracers.

Using Monte Carlo simulations, we modeled a custom designed multifocal gamma probe featuring a 4-segmented collimator and detector. To simulate surgery, a 511 keV radioactive point source was embedded 35 mm below the probe in phantom breast tissue. The source was positioned at various known x, y locations and a 4-channel energy spectrum was recorded. Simulations were repeated 300 times, and the data was split using an 80:20 ratio for training and testing. A 1D convolutional neural network (CNN) was trained to analyze the recorded energy spectra and predict the x, y location of the radioactive source.

The CNN was able to effectively predict the location of the radioactive source from a 4-channel energy spectrum of a multifocal gamma probe. As desired, there was a strong linear relationship ( $R^2 = 0.93$ ) between the true and predicted coordinate locations. The CNN had a small mean prediction error of 2.9 mm and could predict the location of the radioactive source over a large 40x40 mm field of view. The CNN predictions improved the resolution of the multifocal gamma probe by at least 10-fold compared to existing gamma probes. Overall, this work presents a new, real-time localization technique that offers higher resolution and more efficient directional guidance for detecting high-energy radiotracers with a hand-held gamma probe in surgery.

### **Keyword-1**

radioguided surgery

### **Keyword-2**

machine learning

### **Keyword-3**

gamma detector

**Primary author:** WILSON, Sydney (Western University)

**Co-author:** HOLDSWORTH, David (Western University)

**Presenter:** WILSON, Sydney (Western University)

**Session Classification:** DPMB Poster Session & Student Poster Competition (28) | Session d'affiches DPMB et concours d'affiches étudiantes (28)

**Track Classification:** Technical Sessions / Sessions techniques: Physics in Medicine and Biology / Physique en médecine et en biologie (DPMB-DPMB)