

CALOR 2024

Tsukuba

The 20th International Conference on Calorimetry in Particle Physics

Beam Test Results of the Calorimeter Prototype Based on Lead Tungstate Crystal with SiPM Readout

Vladislav Kuskov

ORCID: 0009-0008-2898-3455

CALOR 2024


Tsukuba



Tsukuba, Japan


2024

Introduction



CALOR 2024
Tsukuba



- Good time resolution in calorimetry can provide strong capability to neutral particles identification
 - Hadron PID (neutron/antineutrons, protons/antiprotons etc.)
 - Photon/electron PID
- Lead tungstate crystal (PbWO_4) has demonstrated excellent performance in the experiments in high-energy physics
- Silicon photomultipliers (SiPM) with short rising time of an output signal are capable to provide good time resolution




CALOR 2024
Tsukuba

BEAM TEST CAMPAIGN 2014
PS T10

Experimental layout



Prototype is build of 2x2 crystals matrix:




Readout:




- MPPC S10362-33-025C
- MPPC S10362-33-050C
- MPPC S10362-33-100C
- R7400 PMT



Time Resolution




SiPM/PMT readout at front and rear sides:



Possible explanation (see [ref](#)):

- high reflecting index ($n = 2.2$) \rightarrow low speed of the light propagation
- Simple model shows that the front side is more sensitive to longitudinal shower fluctuations


Similar results with Calvision: [Bob Hirosky](#)



CALOR 2024
Tsukuba

BEAM TEST CAMPAIGNS 2023, 2024
PS T09, SPS H2

Prototype Design



- The calorimeter prototype is build of PbWO_4 crystals size of $22 \times 22 \times 180 \text{ mm}^3$:
 - homogeneous PbWO_4 crystal is served both as scintillator and absorber

Density, g/cm^3	Radiation length, cm	Light yield, % of NaI:Tl	Molière radius, cm	Decay time, ns
8.28	0.89	0.5%	2.2	5-15



- The readout channels consists of Hamamatsu MPPC S14160-6015PS and S14160-6010PS photodetectors:
 - Hybrid SiPM connection: signal in serial, voltage in parallel

1×S14160-6010PS:


- 10 μm pixel pitch
- gain $1.8 \cdot 10^{15}$ (low gain, **LG**)
- for high energy measurements ($E > 10 \text{ GeV}$)

3×S14160-6015PS:



- 15 μm pixel pitch
- gain $3.6 \cdot 10^{15}$ (high gain, **HG**)
- for low energy measurements ($0.5 \text{ GeV} < E < 10 \text{ GeV}$)



Prototype Design




Cluster of 3×3 crystals was equipped with SiPMs



- LG – Low Gain channel for high energy measurements (13-24)
- HG – High Gain channel for low energy measurements (1-12)


Temperature Regime




The light yield (LY) of PbWO_4 crystals significantly depend on the operating temperature ($-2\% \text{LY}/^\circ\text{C}$):

- During the tests, the prototype was thermalized by high-precision cooling plant
- Operating temperature has been set to -24.13°C
- The thermal stability of the prototype is essential during the data-taking period


Temperature distribution
over the whole data-taking period



Temperature distribution per day



Experimental Setup




CA-LOR 2024
Tsukuba

Configuration of the T9 secondary beam at Proton Synchrotron in CERN:

- Primary proton beam of momentum 24 GeV/c + production angle 30 mrad
- Hadron target Be+W (200mm+3mm) for the secondary beam production
- Additional Pb foil converter of 4 mm for e^-/e^+ pair production of momentum $p < 5$ GeV/c


The experimental layout:



- Cherenkov detector XSET048:
CO₂ gas pressure variation → variation of electron signal purity
- Scintillator A (100×5×5 mm³) and scintillator B are used for the trigger system in coincidence connection
- The prototype itself is placed on the DESY table that provide the prototype fine positioning
- Cooling system for the prototype cooling




Experimental Setup





CA-LOR 2024
Tsukuba

The same layout at H2 at SPS:



Trigger System













- All the signals from scintillator A and scintillator B are connected in the coincidence circuit (CC)
- for measuring of electrons $p > 5 \text{ GeV}/c$, signal from Cherenkov detector is added to the CC
- DSn is a clear busy signal from the frontend electronics
- for the time measurements, the reference signal for TDC is also produced by the trigger system

VME frontend electronics:

- CAEN V785 ADC for SiPM amplitude measuring
- CAEN V792 QDC for signal measuring from scintillators
- CAEN V1290 TDC for SiPM time measuring
- CAEN V2718 – V2818 controllers for VME-PCI bridge. DSn is formed by V2718



Energy Calibration



- Beam momentum $p = 2 \text{ GeV}/c$
- Pedestals are subtracted
- Gauss fit of the maximum signal at the right tail of the distribution
- Mean value of Gauss = correspond beam energy at a given channel

→ energy scale for each ADC channel

Energy Resolution







The gravity center of an event is $x = \sum_{i=1}^9 x_i \cdot \frac{E_i}{E_{tot}}$


The range of x_i (y_i) is 1,2,3 (according to chosen matrix 3x3)

To exclude asymmetric clusters the cut on the gravity center has been applied: $R = \sqrt{x^2 + y^2}$


Resulting resolutions are presented below:




Energy Resolution




Combined energy resolution for PS and SPS energies:



Nonlinearity at SPS energies:





EM Leakage Estimation




For the EM shower leakage estimation, an assembly of 3×3 PWO crystals has been installed in front of prototype



→ the prototype measures EM shower leakage from 3×3 absorber in front of it



EM Leakage Estimation




For the EM shower leakage estimation, an assembly of 3×3 PWO crystals has been installed in front of prototype





→ the prototype measures EM shower leakage from 3×3 absorber in front of it

Energy leakage at SPS energies:



Time Resolution




A leading edge discriminator has been exploited for the time and energy measurements


→ The time-energy distribution has a characteristic nonlinear dependence caused by the discriminator threshold

Thus, to increase the accuracy of time resolution calculations, the slewing correction has been applied


A criterion on matching of output signals from edges of central scintillator in front of 3×3 matrix to select for central crystal selection:




Time Resolution



Beam momentum $p = 1 \text{ GeV}/c$, SPS results




Beam momentum $p = 100 \text{ GeV}/c$, SPS results



After applying criterion on central crystal selection, the time resolution reaches value of $\sigma_t < 200 \text{ ps}$ for deposited energies $E > 0.3 \text{ GeV}$


During test beam at SPS, the threshold on discriminator was much higher than one for test beam at PS

Summary



CALOR 2024
Tsukuba

- The prototype has shown good results in terms of time and energy characteristics:
 - time resolution of $\sigma_t < 200$ ps for deposited energies $E > 0.3$ GeV has been achieved
 - nonlinearity of the prototype reach the value of 8% at electron momentum $p = 150$ GeV/c, while EM shower leakage reaches level of 2.5%
 - good energy resolution
- Achieved time and energy resolution could be relevant for the photon physics at low energies and hadron PID (i.e. neutrons/antineutrons) for the future experiments in particle physics




CALOR 2024

Tsukuba


THANK YOU!

Backup slides

Energy Calibration




- The peak in the plot below is taken during the beam stop
- The peak mean position at ≈ 200 MeV corresponds for the muon signals (which penetrate the beam stopper) Energy5mu




The peak position is close to the MIP signal which indicates fine energy calibration of the assembly

Backup slides



Electrical interferences



- During the beam tests, the ADC pedestal positions for each channel have been stable
- Only neglectable pedestal shifts in ± 1 ADC bit have been observed over the whole period of data taking



Energy Resolution Systematic Study



For the energy resolution calculations systematic study has been done:


- Red dots and line stand for mean value of the energy resolution
- Red band stands for systematic uncertainties which comes from gravity center variation, calibration energy variation and Gaus fitting configuration
- Black line stands for the best estimation of measurements
- Blue and magenta lines are reference from previous studies with APDs

Estimated energy resolution:


$$\frac{\sigma E}{E} = \sqrt{\frac{2.344}{E} + \frac{2.344}{\sqrt{E}} + 1.18}$$

Backup slides








Nonlinearity at PS En




- Each dot in the plot below stand for one measurement at the given beam energy (x axis)
- The ration of deposited energy over beam energy (y axis) shows energy leakage form assembly (dots systematically lays under 1)



Energy Resolution at PS energies



Energy Resolution at SPS energies



CALOR 2024
Tsukuba

