MID ISE 2024

Update week 12

M. Verzeroli, M. C. Arena







EP-DT
Detector Technologies



Update

Electrode labelled "ALICE"


- Calibration: 04/06/2024
 - -> next calibration 18/06/2024
- First measurement: 19/02/2024

After purifier and before purifier:

- 320 ml sampling solution

Since no accumulation in the pp collision:


- -> bypassed MFM sensor
- -> flow set from rotameter: 1 ln/h








Excel: calibration






Analysis 2024







Analysis 2018





Comparison 2023 and 2024







Comparison 2023 and 2024

To understand if is related to the change in the WP (~10.3 kV 2018, ~ 9.6 kV mean) -> test can be performed at GIF++ with STD gas mixture:

	Voltage	ABS
Test 1	WP	2.2
Test 2	WP-700	2.2
Test 3	WP	6.9
Test 4	WP-700	6.9