


# Alarm statistics w27-28

G. Rigoletti




EP-DT  
Detector Technologies


# Alarms statistics



<https://gcs-email-alarms-visualization.app.cern.ch/>



# Alarms w27-28



Mattia Piquet (04/07 - 11/07)

## ALICE TRD Xe recuperation dew point HH

→ Alarm received again: [2613](#).

→ Elog with details of the issue: [5925](#)

## ATL MMG stop due to isobutane ratio in mixer

→ Asked PGS team to stabilize isobutane line pressure

## ATLAS PGS R-134a tank LL level alarm

→ Alarm level needs to be updated

## CMS RPC low pressure on recuperated R-134a line