

# Introduction to Data Acquisition System

## Part I

# Outline

What is the goal of this talk?

## Detectors & Readout Electronics

How:

- Our detectors
- The digitizers and A/D converters
- Everything else: Voltage supplies, gas, ...

## Trigger Logic?

## What is happening at T10?


## Organization

- The logbook, good run list, mattermost,
- Where to find the data

# Some Abbreviations

DETECTOR TYPE	NAMES
<b>S0, S1</b>	Scintillator Before and After Cherenkov Detector
<b>XCET0, XCET1</b>	High Pressure and Low Pressure Cherenkov Detector
<b>FS0, FS1</b>	Finger Scintillators
<b>DWC_UP, DWC_DOWN</b>	Delay Wire Chambers
<b>S2, S3</b>	Scintillators (end of the zone)
<b>CAL</b>	Lead Glass Calorimeter


# Our Setup




# Scintillators

3 main steps:


- The scintillator converts the energy from incoming particles into light.
- The PMT captures this light, converts it into electrons, and amplifies the signal.
- The resulting electrical signal is sent to the DAQ system, where it can be recorded and analyzed.



# Scintillators:



# Threshold Cherenkov Detectors (XCET)



- One High – Pressure Threshold Cherenkov Detector (up to approx 15 Bar)
- One Low – Pressure Threshold Cherenkov Detector (up to approx 4.3 Bar)

- Principle:


$$\theta_c \sim \frac{1}{\beta \cdot n(p)}$$

- $\beta = \frac{v}{c}$ ,  $n$  ... refractive index,  
 $p$  ... pressure
- If  $\theta_c$  is above a threshold, then particles emit light as they pass through the gas enclosed in the detector

# DWC

## Delay Wire Chamber

- A **delay wire chamber (DWC)** is a type of particle detector used primarily in high-energy physics to track charged particles. It operates based on the



LEFT



RIGHT

# Calorimeter

- Absorb electromagnetic particles!
- Energy converted to light PMT converts light into electrical pulse
- Integrate pulse to recover the particle energy
- For us: The idea is to give the particles more chance to release photons which might give us a better resolution for the tiny differences in deposited energy compared to what we expect with a 1 cm wide "normal" scintillator.



# Gas Supply

- Ar + CO<sub>2</sub> mixing gas is needed for DWC
- A supply of Argon and Carbon Dioxide (Ar + CO<sub>2</sub>) is provided from a distribution panel behind the beam control room



# Power Supply


## High Voltage Power Supply

- GECO - Software for remotely control HVs:
- For Scintillators: 1500-2500 V negative voltage
- For DWC: 2800-3000 V positive voltage

Custom	Name	I0Set	V0Set	IMon	VMon	Pw	Status	RUp	RDWn	Trip	V1Set	I1Set	SVMMax
00.000	FS0	400.0 uA	1625.0 V	292.0 uA	1625.0 V	On		250 Vps	250 Vps	10.0 sec	0.0 V	300.0 uA	1800 V
00.001	FS1	400.0 uA	1575.0 V	284.5 uA	1575.5 V	On		250 Vps	250 Vps	10.0 sec	0.0 V	300.0 uA	1800 V
00.002	S2	2000.0 uA	2400.0 V	1519.0 uA	2401.0 V	On		250 Vps	250 Vps	10.0 sec	0.0 V	300.0 uA	2500 V
00.003	S3	2000.0 uA	1900.0 V	1136.5 uA	1900.0 V	On		250 Vps	250 Vps	10.0 sec	0.0 V	300.0 uA	2500 V
00.004	CAL17	700.0 uA	1200.0 V	368.5 uA	1200.5 V	On		250 Vps	250 Vps	10.0 sec	0.0 V	300.0 uA	1250 V
04.000	DWC_DUT	100.00 uA	2600.00 V	0.598 uA	2599.85 V	On		250 Vps	250 Vps	10.0 sec	0.00 V	1.00 uA	3000 V
04.001	DWC_UP	100.00 uA	2700.00 V	0.108 uA	2699.73 V	On		250 Vps	250 Vps	10.0 sec	0.00 V	1.00 uA	3000 V
04.002	DWC_DOWN	100.00 uA	2600.00 V	0.096 uA	2599.86 V	On		250 Vps	250 Vps	10.0 sec	0.00 V	1.00 uA	3000 V

# Power Supply

## Low Voltage Power Supply



- Required for the DWCs
- Four units, symmetrically +/- 6V with approx. 50 – 60 mA
- Have to be carefully monitored to detect issues with the DWC before they take damage

# Readout Electronics

## NIM Module


Nuclear Instrumentation Module (NIM)

Modular standard for analog electronic components Rack mount 'bin' holds modules and provides power

Modules:


- Logical and Linear Fan-In/Out
- Discriminator
- Coincidence (logical AND)
- Timer Counter (TDC)

Scaler Disc. Coinc. Voltage Sp.



# That might look like slightly complicated system


- Slightly...



# NIM Modules:

## Discriminator

- A discriminator has two settings;
- a 'Threshold' and a 'Width'
- Noise with a voltage below the Threshold is ignored
- If the input signal crosses the Threshold, a pulse is generated with the desired Width




# NIM Modules

## Coincidence


- Many pulses are not interesting but if two detectors see a signal at the same time, there may be something interesting
- A coincidence is the logical AND of some signals

# Digitization: VME (Versa Module Europa)

- Rack mount 'crate' holds modules, provides power and hosts a data bus (CORBO)
- Single Board Computer (SBC)
- Modules:
  - Scaler/Counter
  - Charge to Digital converter (QDC)
  - Time to Digital converter (TDC)




# Trigger Logic




- Trigger system detects these important moments and sends a signal to the DAQ system, saying, “start recording now!” -just like how the stage lights turn on at the perfect time.

# Trigger Logic

## More Scientific Aspect



# Trigger Logic



# What is DAQ?



- **Data AcQ**quisition is;
- the process of sampling signals –
- that measure real world physical conditions –
- and converting the resulting samples into digital numeric values that can be manipulated by a PC

NIM -> VME -> TDAQ Software




# Busy Logic

- We have to ensure that our active measurements
- are not interrupted by incoming events that arrive a bit later => Block trigger until we are no longer **busy**
- Issuing and clearing the BUSY signal is handled via the CORBO card.
- **We can not record every particle that arrives and that would trigger our detector (at least not in T10)**
- That's why having a scaler is so important, it gives us a less detailed picture about some quantities for(almost) all particles while we take more detailed data (i.e., TDC,




# What is happening at T10?




- Every red line corresponds to a „spill“ of particles received and data being recorded
- There is a structure in there – let's zoom in a bit!

# What is happening at T10




- The spills occur at fairly regular intervals (approx ~ 40 sec)
- According to the built-in detectors at T10, we receive ~ 42k events per spill
- According to our DAQ system, we record ~ 1200-2700 events per spill

# What is happening at T10?



- The regular structure of the spills occurs because multiple users and facilities share the particles accelerated by the PS accelerator
- You can have a real-time view onto this cycle and see when it is your turn to get a spill
- => <https://op-webtools.web.cern.ch/vistar/vistars.php?usr=CPS>

- Example: 2024/09/14 07:02:05
  - The cycle contains 36 „slots“
  - Each „slot“ is about 1.2 seconds long
  - => The cycle repeats every ~ 40 seconds
- seconds
- You can find the slots for T10 under the name „EAST\_N24“
  - The length of the cycle and our position(s) in the cycle can and will change over time



# Organization

Where you can find the data?

In this directory

The screenshot shows the CERNBox web interface. The breadcrumb navigation path is: CERNBox > eos > project > b > bl4s > Technics&Physics > 2024 > Data. The interface includes a search bar, a sidebar with navigation options, and a main content area with a table of items.

İsim ↓	Paylaşımlar	Boyut	Değiştirildi	Aksiyonlar
<input type="checkbox"/> DATA		0 B	0 saniye önce	
<input type="checkbox"/> T10July		11,8 GB	19 gün önce	

2 öge ile 11,8 GB toplamda (0 dosya, 2 klasör)

# Organization

Log Book - 2024 PART II !! : <https://codimd.web.cern.ch/hVW7h8luRrSmj70SsRDMWA?both#>

The screenshot shows a CodiMD web editor interface. The left pane displays the source code in a dark theme, and the right pane shows the rendered page in a light theme.

**Source Code (Left Pane):**


```
1 ---
2 title: Log book BL4S 2024, Part 2
3 tags: BL4S, Logbook, 2024, DWC
4 ---
5
6 # Log Book BL4S 2024, Part 2
7
8 :::warning
9 Useful Links & Pointers
10 - Log book part 1 covering the test beam in July: [HERE]
11   (https://codimd.web.cern.ch/HZC6VD4mQf6wQ1yZPSglQA)
12 -
13 :::
14 ---
15
16 ## 2024/09/12, T10
17
18 ### Patch Panel Connection Table
19
20 **NOTE**: The Patch Panel signal table is expected to
21 change much more often than the actual signal table
22 (see below). We will note updates inline in the comments
23 and only provide a new version of the patch panel
24 connection table if significant enough changes have
25 occurred.
```

**Rendered Page (Right Pane):**

- Page title: Log Book BL4S 2024, Part 2
- Status: CHANGED 2 DAYS AGO, EDITABLE
- Section: Useful Links & Pointers
  - Log book part 1 covering the test beam in July: [HERE](https://codimd.web.cern.ch/HZC6VD4mQf6wQ1yZPSglQA)
- Section: 2024/09/12, T10
- Section: Patch Panel Connection Table
- Note: **NOTE:** The Patch Panel signal table is expected to change much more often than the *actual* signal table (see below). We will note updates inline in the comments and only provide a new version of the patch panel connection table if significant enough changes have occurred.
- Table: **\*\*Patch Panel Connections Table for measurement of S0, S1, XCET040 and XCET043\*\*:**

Connector	Signal	Type	Direction	Comment
B01	S0	NIM	ZONE -> COUNTING	

# Organization Signal



Thank you.  
Questions?