

Alarm statistics w39-40

G. Rigoletti




EP-DT
Detector Technologies


Alarms statistics



<https://gcs-email-alarms-visualization.app.cern.ch/>



Alarms w38-39



Beatrice (26/09-03/10)

- ALICE TRD Mixer injection too long
 - Bottle finished, replaced by Chilo
- CF4 recovery battery almost empty
 - Threshdhold L modified
- HMPID system stop due to gas loss alarm
 - Under investigation
- ATLAS RPC purifier 1 column B offline
 - Followed up by Kacper