

Measurement of thulium-167 radioisotope from alpha-particle induced reaction on holmium for medical application

In the present work, the cross sections of ^{167}Tm radioisotope were measured from α -particles induced reactions on ^{165}Ho . The reaction cross sections were measured in the energy range of 49.6 MeV down to the threshold of production of the thulium-167. The stacked-foil activation technique was used for the irradiation of target whereas the gamma-ray spectrometry was performed using a Ge detector. The measured excitation function for the radioisotope was compared with previous measurements as well as the TENDL-2017 library. We also calculated the thick target yield for the radioisotope. The present results could be useful for medical applications, astrophysical studies, and for the improvements of the nuclear reaction models codes. Specifically, the ^{167}Tm was proposed as an important radioisotope for studying bone and tumor and as skeletal imaging agent [1, 2]. Furthermore, the emission of Auger electrons, low energy γ - and X-rays by the ^{167}Tm and its moderate long half-life (9.25 days) are fitting characteristics for radiotherapy [3,4], making it suitable for the treatment of small tumors weighing 1 mg or less [5].

- [1] R. Chandra, J. Hernberg, *Radiology*. 100, 678 (1971)
- [2] D. Nayak, S. Lahiri, *J. Radioanal. Nuc. Chem.* 242, 423 (1999).
- [3] H. Uusijärvi, P. Bernhardt, *J. Nuc. Med.* 47, 807 (2006).
- [4] F. Tarkanyi, A. Hermanne, *Appl. Rad Isot.* 68, 404 (2010)
- [5] M. Sadeghi, N. Zandi, *J. Radioanal Nuc. Chem.* 291, 731 (2012).

Position

Affiliation

Country

Authors: AHMED RUFAl, Usman (Department of Physics, Umaru Musa Yar'adua University); HABA, H (Nishina Center for Accelerator-Based Science, RIKEN); KHANDAKER, M.U. (Center for Radiation Sciences, School of Health Care and Medical Sciences); OTUKA, N (Department of Nuclear Sciences and Applications, International Atomic Energy Agency)

Presenter: AHMED RUFAl, Usman (Department of Physics, Umaru Musa Yar'adua University)

Session Classification: Parallel Session-III

Track Classification: Medical imaging, security and other applications