



Contribution ID: 275

Type: Poster

## The upside-down structure for X-ray imaging

Typical in-direct X-ray image sensor have used a upper scientillator, which is like CsI(Tl) or Gd<sub>2</sub>O<sub>2</sub>S. In soft X-ray imaging for mammography, most of X-ray Energy is absorbed a surface of scientillator. Thus, MTF and sensitivity of image sensor can be decreased. In our study, we adapted a upside-down structure, which has bottom sided scientillator of image sensor. In MCNP and LightTools simulation, light output and spatial resolution is increased more than 15%. The test is being set in our places and finished before the conference. The test chip have been fabricated in 0.18um 1P3M process.

**Primary author:** Mr MYUNGSOO, Kim (KAIST)

**Co-authors:** KANG, DONG-UK (KAIST); LEE, Daehee (KAIST); Prof. CHO, Gyuseong (KAIST)

**Presenters:** KANG, DONG-UK (KAIST); LEE, Daehee (KAIST); Mr MYUNGSOO, Kim (KAIST)

**Track Classification:** Sensors: 1b) Semiconductor Detectors