



Contribution ID: 253

Type: Oral

## New diamond detector structure and related front-end electronics for TOF application

*Thursday, 5 June 2014 11:40 (20 minutes)*

The results obtained at BTF (Beam Test Facility) of Frascati with 500 MeV electrons working at single electron mode and with cosmic rays have shown a time resolution of the order of 100 ps with a polycrystalline diamond detector of 1.25 mm total thickness and a surface of  $3 \times 3 \text{ mm}^2$  operated at 350 V. To achieve this performance, a new structure of the diamond detector and a dedicated front-end electronics have been developed. The results obtained will be compared with standard mono and polycrystalline diamond detectors. This new structure of the detector together with the dedicated front-end electronics suggest the possibility to realize diamond detectors for MIPs with time resolution of the order of few tens ps.

**Primary authors:** DI CIACCIO, Anna (Universita e INFN Roma Tor Vergata (IT)); PAOLOZZI, Lorenzo (Universita e INFN Roma Tor Vergata (IT)); CARDARELLI, Roberto (Universita e INFN Roma Tor Vergata (IT))

**Presenter:** PAOLOZZI, Lorenzo (Universita e INFN Roma Tor Vergata (IT))

**Session Classification:** I.b Semiconductors

**Track Classification:** Sensors: 1b) Semiconductor Detectors