

# Determination of Trapping Time Constants in Neutron Irradiated Thin Epitaxial Pad Detectors

*Monday 12 November 2007 16:50 (20 minutes)*

Due to the thickness and therefore very short signal length, the classic analysis techniques (ECC, CCM) failed for the determination of trapping times. We presented a method which allows to extract the trapping time constant without a time resolved signal. The method is a combination of simulation and measurement of the charge collection of an irradiated pad detector.

**Primary author:** WEBER, Jens (Universität Dortmund)

**Presenter:** WEBER, Jens (Universität Dortmund)

**Session Classification:** Pad Detector Characterization & Defect Engineering

**Track Classification:** Defect Engineering and Pad Detector Characterization