



GEM based detector for upgrade of the CMS forward muon system

RD51 Mini Week


31.01.2013

Andrey Marinov
on behalf of the
GEMs for CMS Collaboration

Outline



- Introduction:
 - The present CMS Endcap system
 - The case for GEMs at CMS
 - Motivations
- Prototypes construction
 - Large prototypes key points
 - Single-mask GEMs (new technology)
 - Self-Stretching GEMs (new technique)
- The CMS large-size detector
 - System layout,
 - Chamber preparation
 - Detector configurations
 - Gain calibrations
- Test-beam results
 - Timing studies
 - Single-mask results
 - Large-size performance
 - Ageing tests
- Services, integration and installation for the LS1 GE1/1 Demonstrator
- The new CMS GEM lab in bldg.186
- Large-scale GEM production
 - CERN
 - Korea
 - Quality Control
- Electronics system [*μ TCA system*]

Introduction: the present CMS Endcap system



Reduced RPC endcap system
 $|\eta| < 1.6$

GEM upgrade project



The CMS Forward Muon RPC system is equipped with detectors up to $|\eta| < 1.6$, while the high- η region is presently vacant.

➤ Instrument it with a detector technology that can sustain that environment and is suitable for operation at the LHC and its future upgrades.

Motivations

CMS was designed to have a highly Redundant Muon system but we are missing redundancy in the high- η region.

In particular the high- η region needs robust and redundant tracking capability.

Detectors with high resolution would bring additional benefits in Muon HLT, reconstruction and identification.


Improve contribution to Muon Trigger Efficiency.

CMS REGION	Rates [Hz/cm ²]			Charge [C/cm ²]		
	LHC (10 ³⁴ cm ² /s)	High Luminosity LHC (3 10 ³⁴ cm ² /s)	Super LHC (10 ³⁵ cm ² /s)	LHC (10 ³⁴ cm ² /s)	High Luminosity LHC (3 10 ³⁴ cm ² /s)	Super LHC (10 ³⁵ cm ² /s)
Barrel RPC	30	Few 100	~1000 (tbc)	0.05	0.15	~ 1
Endcap RPC <u>$\eta < 1.6$</u>	30	Few 100	~1000 (tbc)	0.05	0.15	~ 1
Endcap RPC <u>$\eta > 1.6$</u>	500 - 1000	Few 1000	Few 10k	(0.05-1)	few C/cm²	Several C/cm²

CMS high-eta region requirements are demanding in terms of rates and integrated charge

Motivations

Muon transverse momentum resolution for different muon system configurations



Instrumenting the inner stations with a double layer of GEM will provide independent pattern recognition and seeding of the track momentum fit.

GEMs, being also a tracking devices, will allow the muon pattern recognition also in partial, and even total absence, of the CSC allowing a direct measurement of the tracking performances using two independent muon systems.

Prototypes construction

Producing large prototypes for CMS

Key points

New single-mask technology

Single-mask technology provides alignment of the GEM foils, while double-mask technology cannot be used for large size foils due to misalignments.

- Small prototypes demonstrated that single-mask GEMs achieves excellent performance.
- Large prototypes confirmed that single-mask technology is mature.

New Stretching technique

The usual thermal stretching is not suitable for the mass production and large-size detectors.

The new single-mask technology

DOUBLE MASK



50 mm polyimide foil, copperclad



photoresist lamination, masking,
exposure and development



metal etching



polyimide etching





metal etching



second masking to define electrodes




metal etching and cleaning



reality

SINGLE MASK




We acknowledge the RD51 Collaboration for the development of the Single-mask technology


Single-mask GEM achieves same performance level as double-mask GEM
Single-mask technology used for large CMS-size prototypes

Stretching techniques

Old thermal stretching



New self-stretching technique



The new self-stretching Technique








The new self-stretching technique has been applied to the full-size CMS detector which will be used for the LS1 CMS GE1/1 Demonstrator





The CMS Full-size NS2 detector prototype for CMS

- GEMs active areas: 990 mm x (220-445) mm
- Single-mask technology
- 1D read-out with (3x10x128) **3184** channels
- HV sector: 35
- Gas mixtures:
 - Ar:CO₂ (70:30; 90:10)
 - Ar:CO₂:CF₄ (45:15:40; 60:20:20)
- Gas flow: ~5 l/h


The CMS full-size layout













Energy Resolution: 19.61%





Energy Resolution: 18.95%







Test beam main results

CMS-RD51 Test Beams (SPS H4 2010)


Test Beam @ RD51 SPS-H4 Setup



CMS full-size detector




GEM Tracking telescope




Oct. 2010 :
20 million events taken with
CMS Proto I

Timing Studies (2010)

Standard GEM Timing Performance




Standard GEM Timing Performance




Custom made HV divider for Standard triple-GEM
Clear effect of gas mixture, and induction and drift field
Timing resolution of 4 ns reached

Single mask performance (2010)



High
achievable
gain!


CMS full-size prototype (SPS@H4 2010)




Data-taking focused on different points along the detector to check uniformity

Preliminary results from the first version of large-size CMS detector showed a very good performance.

CMS full-size residuals





CMS full-size efficiency




CMS full-size prototype (SPS@H2 2011)

At CMS we expect $B_{\perp} \sim 0.6 \text{ T}$ (while $B_{\parallel} \sim 3 \text{ T}$)




Prototypes inside the M1 magnet (side view)



CMS full-size prototype (SPS@H4 2011)


RD51 Scalable Readout System (SRS)




Successful data taking with **APV chip** and **Scalable Readout System (SRS)** developed by RD51 Collaboration, instead of TURBO/VFAT system used before.

Cluster size Proto II

CMS GEM3 X-Hit Cluster size with 35001 good events




CMS GEM3 X-Hit Cluster size with 28033 good events




μ^-/π^- beams

Strip Pitch: 0.573 mm



Strip Pitch: 1.146 mm

Δx distr. Tracker GEM & CMS full-size



CMS full-size prototype:




$$\sigma_x < 110 \mu\text{m}$$

in section with smallest pitch

Ageing test preparation

Motivations

- Ensure a long term operation in CMS
- Understand the effects of the radiation on the materials
- Understand ageing origin (if any) and propose solutions



GEM settings:

Drift: 3kV/cm, Others: 3,5 kV/cm Induction: 5 kV/cm

Gain: $8 \times 10^3 - 10^4$

Gas Mix: Ar/CO₂/CF₄ (45:15:40)


The CMS full-size prototype has been installed in the CERN Gamma Irradiation Facility
The detector performance will be monitored along with environmental/gas variables

Services, integration, installation for the LS1 GE1/1 Demonstrator






Define the space available in the YE-1 near side Towers

For Super Chambers

- Using the **RE-1/1/20** and **RE-1/1/21** chamber slots on far side
- Alternatively **RE-1/1/35** et **36** on near side




Integration studies



LS1 GE1/1 – Powering Schema

USC55

UXC55




Gas infrastructures already existing

Chamber distribution (UXC)

One manifold (12 supply and return channels) is available in each distribution racks.

At the moment the controls are integrated in the RPC gas system. We need to decouple the two systems and re-calibrate the flow-meters.




R. Guida

Cooling Request

- The heating power of GE1/1 Demonstrator will be 120W in total for the 2 super chambers.
- It will give negligible impact to the YE1 cooling system.

GE1/1 Cooling Manifold



Use the existing cooling infrastructure in the YE-1 Nose


Taking the cooling from the RBX

The New CMS GEM Lab. In Bldg.186 @ CERN

TIF – bldg. 186



Plan of the working space



Part of the Working area




- The gas installation is in progress. Should be delivered by the end of March.
- Lab infrastructure is available.
- We can start moving soon from RD51 lab to b.186.

GEM production

GEM Production at/outside CERN

The Goal is to setup a production line being able to face most of the future requests for large GEM productions (from 1 piece to few hundreds, size up to 2m x 0.6m)

New Machine procurement and installation⁽²⁰¹¹⁾ and new building⁽²⁰¹³⁾



CERN Building 107
Basis of Design

@CERN

@Korea

New Flex (South Korea near Seoul) has produced 8x8cm² and 10x10cm² GEM; successfully tested at RD51 lab


- Initial contact in 2008; new visit in June 2011
- Full technology transfer done; large size foils to try gradually...




RD51 Mini Week Jan. 2013

Characterization of Korean GEM


Gain Calibration [Ar/CO₂ (70:30)]




Triple GEM configuration stack




Divider performance



Iron 55 : Spectrum and Count Rate





Quality Control (QC)



Quality Control procedures:

1. *QC-Drift Electrode*
2. *QC-GEM foils*
3. *QC frames (mechanical tolerances)*
4. *QC readout PCB*
5. *QC Gain uniformity test*
6. *QC leak test*




The electronics system

The μ TCA development

Design a system that is:

Flexible in terms of detector segmentation.
Uses generic design work as much as possible.



A possible off detector partition:

1 GLIB = one phi segment ie. $10^\circ/3$

1 μ TCA crate = 30° degrees in phi

12 μ TCA crates = 360° (24 μ TCA crates for both endcaps)

Possible FE chips...

VFAT3:

Front-end with programmable shaping time.

Internal calibration.

Binary memory

Interface directly to GBT @ 320Mbps.

Designed for high rate $10\text{kHz}/\text{cm}^2$

(depending on segmentation)

GdSP:


Similar to VFAT3 except has an ADC / channel instead of a comparator.

Internal DSP allows subtraction of background artifacts enabling a clean signal discrimination.

Centre of gravity a possibility to achieve a finer pitch resolution


Measurements to optimize the electronics

Capacity measurement large-size detector




The measurement performed on the readout board and on the GEM signal output will drive the development of the new on- and off-electronics

Inter-strip capacity



Strip pitch as a function of the detectors length



Conclusions


Scope: instrument the vacant high- η region with detectors suitable for high rate, capable of tracking and triggering

- Detector development:
 - New industrial technologies and new cost-effective assembly techniques developed for large-size GEM foils (*large size foils production with no spacers in active area, reduced assembly time, no gluing and no soldering required, possible to re-detector*)
 - High rate, rad. hard triple-GEM technology developed and demonstrated for large area detectors
- Improvements in muon tracking and triggering promising
 - Substantial increase in the acceptance
 - Redundancy and robustness in high- η
 - Lot of interest from trigger and physics groups
- Electronics development: underway
- Integration and services in CMS: studied in sufficient detail
- Preparation of the GE1/1 Demonstrator is in mature stage
- Large participation: currently 39 institutes with 182 collaborators

We gratefully acknowledge the RD51 Collaboration for its strong support of our detector construction, testing and data-taking and for the many fruitful discussions


Thanks to ...

The GEMs for CMS Collaboration:
182 collaborators, 39 institutions, ... countries




Backup Slides

Short Project History (2009-2010)



GE1/1 Proto I has been designed, built and successfully tested in only 1 year !

Short Project History (2011)



*The detector **GE11_II** with enhanced performance has been designed, built and successfully tested in a very strong magnetic field, a small 30x30 detector with a different stretching technique is nearly ready for testing*


Estimated Particle Rates

RPC Region	Rates Hz/cm ² LHC (10 ³⁴ cm ² /s)	High Luminosity LHC	SLHC ?? (10 ³⁵ cm ² /s)?
RB	30	Few 100	kHz
RE 1, 2, 3,4 $\eta < 1.6$	30	Few 100	kHz
Expected Charge in 10 years	0.05 C/cm ²	0.15 C/cm ²	~ C/cm ²
RE 1,2,3,4 $\eta > 1.6$	500Hz ~ kHz	Few kHz	Few 10s kHz
Total Expected Charge in 10 years	(0.05- 1) C/cm ²	few C/cm ²	Few 10s C/cm ²

GEM Rate Capability

Measured with an X-ray (5.9 keV) tube, Ar/CO₂/CF₄ (60/20/20)

Gain of about 2×10^4



CMS high- η - maximum rate


LHCb

Triple GEM Ageing test

Gain of 2×10^4

Total integrated charge of 13 C/cm^2 is expected in 10 years of operation in LHCb



50 MHz/cm^2 X-rays, in 10 days a total charge of 20 C/cm^2 was integrated



Small Prototypes (2009-2010)

- 10x10 cm² triple-GEMs, 1D or 2D readout, 128 or 256 channels :


- ☐ Standard double-mask triple-GEM - "Timing GEM"
- ☐ Single-mask triple-GEM
- ☐ "Honeycomb" triple-GEM




Characterization of GEM Detectors for Application in the CMS Muon Detection System
2010 IEEE Nucl. Sci. Symp. Conf. Rec. 1416-1422; RD51 Note 2010-005; arXiv:1012.3675v1 [physics.ins-det]

Double-Mask vs. Single-Mask GEMs


Base material = Polyimide 50 μ m + 5 μ m copper cladding on both sides



Achieved 40x40cm²




Achieved 200x60cm²




Single vs Double-Mask GEM

Standard GEM Efficiency



Single Mask GEM performance





Single-mask GEM reaches similar performance level as double-mask GEM

➤ Single-mask technique used for large CMS-size prototypes

“Honeycomb” GEM Imaging




Honeycomb GEM config 2




- Insert honeycomb spacers between GEM foils to avoid foil stretching
- Efficiency ~70%; geometrical factor due to spacers roughly estimated at ~65%

Similar performance level observed as for standard GEMs, but local efficiency losses due to spacers (similar effect as seen in CMS RPC)

CMS Proto I Mock-up & CMS Proto I




Readout PCB Proto I



128 channels per VFAT connector

256 strips for each eta partition




0.8mm pitch

1.6mm pitch

PCB thickness = 3mm


Optimized HV Divider

The HV divider is made by HV SMD Resistors and has built-in RC Filter which cuts the intermediate frequency from the CAEN 1527 power supply (GEM detectors are sensible to HV power supply fluctuation).






Resistors value have been chosen according to the timing GEM performance in order to reproduce on the full-scale prototype the excellent behavior of the optimized timing GEM.

Use of ZIF sockets to connect to GEM terminals




SRS System (APV)



The system was successfully adopted with the APV chip during the September 2011 test beam

The RD51-SRS system SPARE

General purpose multi-channel readout solution for a wide range of detector types, detector complexities, and different experimental environments.



Product of the RD51 Collaboration for the Development of MPGD Technologies



- Scalable - size and applications
- Only point-to-point links. No busses
- Star topology
- Allows the use of different front-ends
- Can integrate different sub-detectors DAQ in the same system
- Cost effective
- Use of cost-effective components from high-volume markets (eg. HDMI cables, PCIe connectors, Cat5/6 UTP cables, ...)

**Production of SRS base components externalized.
Purchase via CERN store.**

Full VFAT2 SRS system (HW & SW) may be ready for Aug-Sep. 2012.


Possible upgrade path with xTCA, optical, GBT, ...

A lot of interest from RD51 for VFAT2/VFAT3/GdSP.


TURBO (VFAT) CDARE

The VFAT(TOTEM) is a digital on/off chip for tracking and triggering with an adjustable threshold for each of the 128 channels; it uses 0.25 μ m CMOS technology and its trigger function provides programmable “fast OR” information based on the region of the sensor hit.

Turbo board layout



For prototype testing we used electronics developed by INFN (Siena and Pisa), based on the TOTEM VFAT chip.



Prototypes construction (2009-2012)

REARRANGE! prototypes

- “CMS Timing GEM”: Standard double-mask; 10x10cm²; 1D readout; (3/2/2/2); 256 channels
- “Single-Mask GEM”: Single-mask; 10x10cm²; 2D readout; (3/2/2/2); 512 channels
- “Honeycomb GEM”: Standard double-mask; 10x10cm²; 1D readout; (3/2/2/2); 256 channels

Characterization of GEM Detectors for Application in the CMS Muon Detection System
2010 IEEE Nucl. Sci. Symp. Conf. Rec. 1416-1422; RD51 Note 2010-005

- CMS Proto III : Single-mask; 10x10cm²; NS2; (3/1/2/1); 256 channels
- CMS Proto IV : Single-mask; 30x30cm²; NS2; (3/1/2/1); 256 channels
- Korean I : Double-mask; 7x7cm²; (3/2/2/2); 256 channels

Full-size prototypes

- CMS Proto I: Single-mask; CMS FULL-SIZE; 1D readout; (3/2/2/2); 1024 channels

Construction of the first full-size GEM-based Prototype for the CMS High- η Muon System
2010 IEEE Nucl. Sci. Symp. Conf. Rec. 1909-1913; RD51 Note 2010-008

- CMS Proto II: Single-mask; CMS FULL-SIZE; 1D readout; (3/1/2/1); 3072 channels
- CMS Proto V, VI: Single-mask; CMS FULL-SIZE; 1D readout; NS2; (3/1/2/1) ~3072 channels