



Contribution ID: 247

Type: **Poster presentation (105min)**

Check-Valve with low Cracking Pressure for Gas Recovery Systems

Thursday 10 July 2014 10:30 (2h 15m)

To prevent backflow of boil-off gas into a cryostat or any other cryogenic system, a check valve with a low cracking pressure may be installed. This check valve should be placed in the boil-off gas line at ambient temperature. Design considerations and applications of such low-cracking- pressure check valves will be described and presented.

Author: HOLDENER, Fridolin (WEKA AG)

Co-author: Mr ERNI, Pascal (WEKA AG)

Presenter: HOLDENER, Fridolin (WEKA AG)

Session Classification: Thu-Mo-Posters Session 3.2

Track Classification: C-05: Cryostat technology