

New CCE results with microstrip detectors made on various substrates

Tuesday, 3 June 2008 13:40 (20 minutes)

The charge collection efficiency of μ -strip sensors made with the RD50 mask set with various silicon substrates (n and p FZ and MCz) have been compared to different high doses of neutron and protons, well in the range of the anticipated fluences in SLHC.

Primary author: CASSE, Gianluigi (Department of Physics)

Co-authors: AFFOLDER, Antoni (Department of Physics); WORMALD, Mike (Department of Physics)

Presenter: CASSE, Gianluigi (Department of Physics)

Session Classification: Full Detector Systems