



Contribution ID: 94

Type: **contributed talk**

## Magnetometry standard for the muon $g-2$ experiment

*Tuesday, 8 April 2014 14:30 (15 minutes)*

A key requirement of the  $g-2$  experiment, aiming to measure the muon anomalous magnetic dipole moment to 0.14ppm, is to monitor the magnitude of the magnetic field around the storage ring to 0.07ppm. This will be done with several hundred proton NMR magnetometers. These must all be calibrated against a standard magnetometer, to account for the  $\sim 0.1$ ppm shifts of individual probes due to the magnetic properties of the materials. We are developing a  $^3\text{He}$  magnetometer to provide an independent check of the standard calibration probe used for the Brookhaven  $g-2$  experiment, which was based on a spherical water sample.

**Primary author:** Dr HENRY, Sam (University of Oxford)

**Presenter:** Dr HENRY, Sam (University of Oxford)

**Session Classification:** Parallel 1E

**Track Classification:** Flavour Physics and Beyond