Contribution ID: 142

Type: Poster

A Custom Real-Time Ultrasonic Instrument for Simultaneous Mixture and Flow Analysis of Binary Gases in the CERN ATLAS Experiment

Custom ultrasonic instruments have been developed for simultaneous monitoring of binary gas mixture and flow in the ATLAS Inner Detector. Sound transit times are measured in opposite directions in flowing gas. Flow rate and sound velocity are respectively calculated from their difference and average. Gas composition is evaluated in real-time by comparison with a sound velocity/composition database, based on the direct dependence of sound velocity on component concentrations in a mixture at known temperature and pressure. Five devices are integrated into the ATLAS Detector Control System. Three instruments monitor coolant leaks into N2 envelopes of the SCT and pixel detectors. Resolutions better than $\pm 5 \times 10^{-5}$ and $\pm 10^{-4}$ are respectively seen for C3F8 and CO2 leak concentrations in N2. A fourth instrument detects sub-percent levels of air ingress into the C3F8 condenser of the new thermosiphon coolant recirculator. Following extensive studies a fifth instrument was built as an angled sound path flowmeter to measure the high returning C3F8 vapour flux (~1.2 kg/s). A precision of < 2% F.S. for flows up to 15 m/s was demonstrated. This device can also monitor C3F8 and C2F6 concentrations to better than $\pm 3 \times 10^{-3}$. These blends allow for lower temperature silicon tracker operation.

The instrument has many potential applications where continuous binary gas composition measurement is required, including hydrocarbon and anaesthetic gas mixtures.

Primary author: O'ROURKE, Abigail Alexandra (Deutsches Elektronen-Synchrotron Hamburg and Zeuthen (DE))

Co-authors: BITADZE, Alexander (University of Manchester (GB)); ROZANOV, Alexandre (CPPM, Aix-Marseille Université, CNRS/IN2P3 (FR)); PEARSON, Ben (University of Oklahoma (US)); DI GIROLAMO, Beniamino (CERN); DETERRE, Cecile (Deutsches Elektronen-Synchrotron Hamburg and Zeuthen (DE)); ROSSI, Cecilia (Universita e INFN Genova (IT)); DEGEORGE, Cyril (Indiana University (US)); ROBINSON, Dave (University of Cambridge (GB)); LOMBARD, Didier (CERN); FAVRE, Gilles (CERN); HALLEWELL, Gregory (Centre National de la Recherche Scientifique (FR)); AHMED, Hasib (University of Oklahoma (US)); Mr YOUNG, Joel (University of Oklahoma); NAGAI, Koichi (University of Oxford (GB)); ZWALINSKI, Lukasz (CERN); DOUBEK, Martin (Czech Technical University (CZ)); BATTISTIN, Michele (CERN); STRAUSS, Mike (University of Oklahoma (US)); ALHROOB, Muhammad (University of Oklahoma (US)); CRESPO-LOPEZ, Olivier (CERN); BONNEAU, Pierre (CERN); BATES, Richard (Department of Physics and Astronomy-University of Glasgow); VAGLIO, Romain (Inst. Universitare de Technologie d'Annecy (FR)); BOYD, Rusty (University of Oklahoma (US)); KATUNIN, Sergei (B.P. Konstantinov Petersburg Nuclear Physics Institute - PNPI (); BERRY, Stephane (CERN); MCMAHON, Stephen (STFC - Rutherford Appleton Lab. (GB)); VACEK, Vic (Czech Technical University (CZ))

Presenter: O'ROURKE, Abigail Alexandra (Deutsches Elektronen-Synchrotron Hamburg and Zeuthen (DE))

Track Classification: Gaseous Detectors