

Imaging for Online Control of Particle Therapy

Thomas Bortfeld and Joost Verburg

Massachusetts General Hospital, Boston

tbortfeld@mgh.harvard.edu


Supported by:

MGH-NCI: Therapy Imaging Program (TIP)

ICTR-PHE 2016


Imaging particle beams for cancer treatment

Jeremy C. Polf and Katia Parodi



Physics Today 68(10), 28 (2015)

Example: Para-spinal



Proton Beam




MASSACHUSETTS
GENERAL HOSPITAL

RADIATION ONCOLOGY


Example: Prostate

Have to use lateral beams!

IMRT



Protons




From: Trofimov et al., IJROBP 69(2):444-453, 2007

Example: Lung

Breathing and range uncertainties


State of the art (2016)



With all margins added, proton plans look like photon plans.

Example: Pediatric, medulloblastoma

Thecal-sac-only irradiation?



Range uncertainties and clinical impact: Range margin “recipe”

$$1.5 \cdot \sqrt{2\%^2 + 1\%^2} + 1 \text{ mm} \gg 3.5\% + 1 \text{ mm}$$

standard
deviations

Hounsfield
Units (HU)

HU
conversion

setup
etc.


Michael Goitein 1980




MASSACHUSETTS
GENERAL HOSPITAL

RADIATION ONCOLOGY

In vivo range verification



Imaging for online control of particle therapy: A control loop approach




The ideal range “sensor” – requirements:

- Detect range with ~ 1 mm accuracy
- Short response time
- Low dose (< 1 Gy)
- In the patient
- Relative to patient anatomy




Prompt gamma imaging (PGI) – principle

- Measure secondary **prompt gamma radiation** from nuclear reactions of protons with tissue




Prompt gamma imaging (PGI) – principle



Min et al., Appl. Phys. Lett. 89, 183517 (2006)
Stichelbaut and Jongen (IBA), PTCOG 2003

Potential advantage of prompt gamma range imaging over PET



Clinical implementation slit camera

In vivo range verification: Prompt γ -ray imaging

- Prompt γ -ray imaging (PGI):
Clinical implementation of IBA slit camera



- 1st prototype in Dresden since 10/2014
- 1st clinical application 08/2015
- Evaluation of the benefit under clinical conditions
- **Final Goal:** Reduction of range uncertainties and irradiation of normal tissues

Richter et al. Radioth Oncol 2016, ~~Thu 11:45~~




HPRT,
In vivo dosimetry,
RTTs, clinicians



Improvements and extensions of PGI

- Compton camera for “better” geometry




Improvements and extensions of PGI

- Secondary proton imaging for carbon ion therapy

Prompt γ -ray spectroscopy (PGS)

Unique method to determine the range of proton beams:
in vivo **spectroscopy** of proton-induced **prompt γ -rays**



Prompt gamma detector


Prototype design

7.6 cm $\text{LaBr}_3(\text{Ce})$
scintillator actively
shielded with 4 BGO
scintillators


Photomultipliers for
detector readout

Tungsten collimator


- High energy resolution for spectroscopy
- Fast timing to separate proton- and neutron-induced reactions



Experimental setup




Energy- and time-resolved γ detection



165 MeV proton beam in water, 20 mm upstream of end-of-range


Detection of range shifts





γ -ray energy spectrum

165 MeV proton beam stopped in water (oxygen) and polyethylene (carbon)

Proton-induced component averaged over last 30 mm of range



Discrete γ -ray emissions



Range verification method

Optimization parameters

d : Water equivalent depth at detector position

ρ_O, ρ_C : $^{16}\text{O}/^{12}\text{C}$ concentrations

$$\operatorname{argmin}_{d, \rho_O, \rho_C} \left(\sum_{b \in B} \sum_{l \in L} h \left(\underbrace{\rho_O f_{b,l}^O(d)}_{\text{Reactions with } ^{16}\text{O}} + \underbrace{\rho_C f_{b,l}^C(d)}_{\text{Reactions with } ^{12}\text{C}} - \mathbf{m}_{b,l} \right)^2 \right)$$

Pencil-beams and gamma lines

Loss function

Reactions with ^{16}O

Reactions with ^{12}C

Prompt gamma model


Measurement data



Range verification of pencil-beams

- Pencil-beam delivered at one position with 5 different energies (R_{80} 15.54 – 17.59 g cm⁻²)
- 2.3 mm / 5.2 mm water equivalent shifters in front of phantom to simulate range error

End-of-range in water
(89% oxygen)



End-of-range in solid water
(21% oxygen, 70% carbon)




Combined proton range verification and elemental composition

	Detected depth (mm)	Range shift (mm)	Detected ^{16}O (g cm^{-3})	Detected ^{12}C (g cm^{-3})
Water Detector at 156.5 mm (water equivalent), 0.89 g cm^{-3} oxygen				
No range shift	156.3 ± 1.3	0.1 ± 1.3	0.94 ± 0.07	0.05 ± 0.07
2.3 mm shifter	158.6 ± 1.0	-2.2 ± 1.0	0.90 ± 0.07	0.02 ± 0.06
5.2 mm shifter	161.6 ± 1.0	-5.1 ± 1.0	0.88 ± 0.07	0.04 ± 0.06
Solid Water Detector at 157.1 mm (water equivalent), 0.21 g cm^{-3} oxygen, 0.70 g cm^{-3} carbon				
No range shift	157.1 ± 1.4	0.0 ± 1.4	0.17 ± 0.04	0.70 ± 0.07
2.3 mm shifter	159.3 ± 1.4	-2.3 ± 1.4	0.18 ± 0.03	0.70 ± 0.07
5.2 mm shifter	162.4 ± 1.3	-5.4 ± 1.3	0.18 ± 0.04	0.69 ± 0.08

Pencil-beams delivering 5×10^8 protons, uncertainties $\pm 1\sigma$,
 90 repeat measurements for determination of σ


Planned clinical implementation

- Create **nuclear reaction models** based on treatment plan
- **Prompt γ -ray spectroscopy** at fixed positions near distal end of tumor
- Measurements **fitted to models**, simultaneously optimizing unknown parameters:
 - **Absolute proton range**
 - Elemental composition of irradiated tissue



Imaging particle beams for cancer treatment

Jeremy C. Polf and Katia Parodi



Physics Today 68(10), 28 (2015)

PGI/PGS requires 3D anatomical imaging



MASSACHUSETTS
GENERAL HOSPITAL

RADIATION ONCOLOGY

Revival of acoustic range measurements

NUCLEAR INSTRUMENTS AND METHODS 161 (1979) 203-217 ; © NORT

Radiation Oncology Investigations 3:42-45 (1995)

Ionoacoustic characterization of the proton Bragg peak with submillimeter accuracy

W. Assmann^{a)}

Department for Medical Physics, Ludwig-Maximilians-Universität München, Am Coulombwall 1, Garching 85748, Germany

S. Kellnberger

Institute for Biological and Medical Imaging, Technische Universität München and Helmholtz Zentrum München, Ingolstädter Landstrasse 1, Neuherberg 85764, Germany

S. Reinhardt, S. Lehrack, A. Edlich, and P. G. Thirolf

Department for Medical Physics, Ludwig-Maximilians-Universität München, Am Coulombwall 1, Garching 85748, Germany

M. Moser and G. Dollinger

Institute for Applied Physics and Measurement Technology, Universität der Bundeswehr, Werner-Heisenberg-Weg 39, Neubiberg 85577, Germany

M. Omar and V. Ntziachristos



Institute for Biological and Medical Imaging, Technische Universität München and Helmholtz Zentrum München, Ingolstädter Landstrasse 1, Neuherberg 85764, Germany

K. Parodi

Department for Medical Physics, Ludwig-Maximilians-Universität München, Am Coulombwall 1, Garching 85748, Germany

(Received 6 August 2014; revised 3 December 2014; accepted for publication 7 December 2014; published 9 January 2015)

Acoustic proton range measurement IBA cyclotron (UPenn)



The ideal range “sensor” – requirements:


- Detect range with ~ 1 mm accuracy
- Short response time
- Low dose (< 1 Gy)
- In the patient
- Relative to patient anatomy




Comparison of techniques

	PET	Prompt γ	Acoustic	MRI
1 mm accuracy	(✓)	(✓)	(X)	✓
Response time	minutes	pico seconds	milli seconds	days
Low dose	1 Gy	1 Gy	1 Gy	5-20 Gy
In vivo	✓	(✓)	(X) one in 1995	(✓) bone, liver
Patient anatomy	(✓) (PET/CT)	X	X	✓

Imaging for online control of particle therapy: A control loop approach




MRI scan before and after treatment



Krejcarek, Yock, et al.
IJROBP 2007;68(3):646-649

Biology: Bone marrow changes Visible in MRI

planned 50% isodose
MRI 50% isodose




0 1 2 3cm

Pt. 1, 36 Gy (RBE)




0 1 2 3cm

Pt. 5, 50.4 Gy (RBE)



0 1 2 3cm

Pt. 6, 36 Gy (RBE)





0 1 2 3cm

Pt. 10, 36 Gy (RBE)

Biology: Focal liver reaction

Visible in contrast enhanced MRI


MRI contrast agent : Gd-EOB-DTPA
("Eovist", "Primovist")




See also Y. Yuan et al.
Radioth Oncol 106:378-382, 2013

Patient #5

Post 3fx



Post 5fx



Difference Image (Post 5fx - Post 3fx)

