

Tuning of a 4D ML reconstruction strategy for PET-based treatment verification in ion beam radiotherapy

Elisabetta De Bernardi¹, Katia Parodi², Rosalinda Ricotti³,
Marco Riboldi⁴, Guido Baroni⁴, Chiara Gianoli^{2,5}

¹ Medicine and Surgery Department – Tecnomed Foundation, University of Milano-Bicocca, Milan, Italy

² Department of Experimental Physics - Medical Physics and Department of Radiation Oncology, Ludwig Maximilian University of Munich, Germany; Department of Radiation Oncology, Heidelberg University Hospital, Germany

³ Department of Radiation Oncology, European Institute of Oncology, Milano, Italy

⁴ Dipartimento di Elettronica, Informazione e Bioingegneria (DEIB), Politecnico di Milano, Italy


⁵ Department of Radiation Oncology, Ludwig Maximilian University of Munich, Germany



PET-based treatment verification in ion beam therapy

ICTR-PHE
2016

- Detect possible mismatches with respect to the treatment planning (i.e., patient positioning, intra- and inter-fractional anatomical changes)
- Imaging of secondary radiation of the ion beam treatment
- PET-CT comparison in terms of activity-anatomy




- Poor image quality of the Measured PET (i.e., time delay and washout in *off-line PET*) (Kurz et al. 2015 Med Phys)
- Quantification of the mismatch is performed as *range difference* along the beam direction (1D) (Bauer et al. 2013 Radiother Oncol)



The idea

- Interpret the Expected PET and the Measured PET as different motion states of a 4D PET dataset
- Adapt a 4D Maximum Likelihood (ML) reconstruction algorithm developed for 4D PET imaging to PET-based treatment verification



De Bernardi et al., Med Phys, 2015

*Blume et al., IEEE Trans Med Imag, 2010;
Blume et al., Phys Med Biol, 2012*

Extensive testing on analytical phantom

Sensitivity to:


- Mismatch

Robustness against:

- Activity complexity
- Washout inaccuracy
- Noise

Limitations:

- ↑ Noise ↑ Smoothing





On clinical datasets

Algorithm input


Expected PET
sinogram

Measured PET
sinogram




Algorithm output


Expected PET



Reference
frame



Measured PET



Motion field
Reference frame \rightarrow Expected PET


Motion field
Reference frame \rightarrow Measured PET



On clinical datasets

ICTR-PHE
2016


Algorithm output




The aim of the work

Optimize the framework of the algorithm for clinical datasets by:

- Reducing the number of unknown variables (deformation fields)




- Investigating two different initializations for the reference frame:
 - Uniform activity
 - Expected PET (as prior knowledge)




The aim of the work

- Original method (Method 1): reference frame is an intermediate state between the Expected PET and the Measured PET




The aim of the work

- Method 2: reference frame is the Expected PET frame




The aim of the work

- Method 3: reference frame is the Measured PET frame



Analytical phantom

- Expected PET: combination of six Gaussian spots
- Measured PET: deformation of the Expected PET according to known motion fields




- Washout inaccuracy (66% activity scaling for 2 spots)
- Poisson noise simulation (maximum 2 counts/voxel) + LSO random noise




Analytical phantom: convergence assessment

Similarity vs. Measured PET (Gold Standard) and vs. Expected PET



Analytical phantom: reconstructed images

ICTR-PHE
2016



Trans-axial

Coronal

Sagittal

Analytical phantom: motion fields


ICTR-PHE
2016

Trans-axial

Gold
Standard

Method 2
Iniz. Expected PET | Iniz. Unif.

Method 3
Iniz. Expected PET | Iniz. Unif.



A-P

L-L

C-C



Analytical phantom: motion fields

Coronal








Gold
Standard

Method 2

Iniz. Expected | PET Iniz. Unif.

Method 3


Iniz. Expected | PET Iniz. Unif.



A-P

L-L

C-C



Sagittal




Gold
Standard

Method 2

Iniz. Expected | PET Iniz. Unif.

Method 3


Iniz. Expected | PET Iniz. Unif.



A-P

L-L

C-C



Outcomes


ICTR-PHE
2016

- Application of method 3 with uniform initialization on clinical datasets
- Improvement of noise mitigation technologies and methodologies to better tackle the LSO random noise
 - Time of Flight reconstruction
 - Targeted reconstruction for computational optimization (GE patent 2010)

*Gianoli et al. ICTR-PHE 2016
Proffered papers – Emerging
technologies (Wednesday)*



Outlook on clinical dataset



Discussions

ICTR-PHE
2016

- Refinement of the simulation of the washout inaccuracies (not only spot scaling but generation of pattern of washout inaccuracies)
- Inclusion of Jacobian-based as “washout compensation index” on the motion field
- Disentangle washout compensation from the actual mismatch



Acknowledgments

ICTR-PHE
2016

Heidelberg Ion beam Therapy Center for clinical data

- Jürgen Debus
- Julia Bauer
- Christopher Kurz

