



Contribution ID: 82

Type: **Contributed Talk**

An active pixel sensor and microelectrode array for retinal stimulation

Wednesday 14 September 2005 09:30 (15 minutes)

Degenerative photoreceptor diseases, such as age-related macular degeneration and retinitis pigmentosa, are the most common causes of blindness in the western world. A

potential cure is to use a microelectronic retinal prosthesis to provide electrical stimulation to the remaining healthy retinal cells. Due to the success of cochlear implants in restoring hearing to the profoundly deaf this is becoming a widespread research area.

Here we describe a system capable of detecting the visual scene and translating the

image into a train of electrical pulses that stimulates live retinal tissue. This system

requires a position-sensitive detector to act as a smart retinal chip. We have developed

a CMOS active pixel sensor with a 10 by 10 array of 100 μ m pixels. Each pixel contains

a photodiode and on-pixel circuitry that translates the intensity of the incoming light into a certain frequency of output voltage pulses.

The outputs of the pixels are connected to a biocompatible microelectrode array, which

makes contact with the retinal cells. This electrode array has 74-electrodes spaced at

60 μ m and is fabricated on a flexible polyimide substrate that is only 20 μ m thick. Each

electrode is a platinum disc of diameter 5 μ m, which forms the interface to the retinal cells.

We have verified the electrical contact between the electrodes and the retina by record-

ing the response of the output (ganglion) cells to light. A typical signal to noise ratio

of 30:1 has been achieved with an overall system noise of 5 μ V rms. Retinal cells of

only 10 μ m in diameter have been stimulated by a range of voltages from 0.1 to 2V

with pulse durations of 0.1-1 μ s. The effects of charge spreading within the retina have

been studied and optimal stimulation parameters determined. This system permits the simultaneous stimulation and recording of neural activity on the retina and allows the image processing that occurs within the retina to be studied.

Primary author: Dr MATHIESON, Keith (University of Glasgow)

Presenter: Dr MATHIESON, Keith (University of Glasgow)

Session Classification: S7 : Novel Photon Detection Systems

Track Classification: New Techniques for Positron Emission Tomography