



Contribution ID: 327

Type: **Oral Presentation**

The Phase-1 Upgrade of the CMS Pixel Detector (12' + 3')

Thursday 4 August 2016 17:45 (15 minutes)

The innermost layers of the CMS tracker are built out of pixel detectors arranged in three barrel layers (BPIX) and two forward disks in each endcap (FPIX). The original CMS detector was designed for the nominal instantaneous LHC luminosity of $1 \times 10^{34} \text{ cm}^{-2} \text{ s}^{-1}$. Under the conditions expected in the coming years, which will see an increase of a factor two of the instantaneous luminosity, the CMS pixel detector will see a dynamic inefficiency caused by data losses due to buffer overflows. For with full eff this reason the CMS Collaboration has been building a replacement pixel detector which is scheduled for installation in an extended end of year shutdown during Winter 2016/2017.

The Phase I upgrade of the CMS pixel detector will operate at full efficiency at an instantaneous luminosity of $1 \times 10^{34} \text{ cm}^{-2} \text{ s}^{-1}$ with increased detector acceptance and additional redundancy for the tracking, while at the same time reducing the material budget. These goals are achieved using a new readout chip and modified powering and readout schemes, one additional tracking layer both in the barrel and in the disks, and new detector supports including a CO₂ based evaporative cooling system, that contribute to the reduction of the material in the tracking volume.

This contribution will review the design and technological choices of the Phase I detector, and discuss the status of the construction of the detector and the performance of its components as measured in test beam and system tests. The challenges and difficulties encountered during the construction will also be discussed, as well as the lessons learned for future upgrades.

Presenter: MERKEL, Petra (Fermi National Accelerator Lab. (US))**Session Classification:** Detector: R&D and Performance**Track Classification:** Detector: R&D and Performance