



Contribution ID: 1211

Type: Poster Presentation of 1h45m

Design and Preliminary Test Results of the 60T Pulsed High Magnetic Field Facility at Xi'an Jiaotong University

Monday 28 August 2017 13:15 (1h 45m)

A 60T pulsed high magnetic field facility is designed for the study of magnetic material at Xi'an Jiaotong University. The pulsed high magnetic field facility is energized by two 1.2MJ/3.84mF capacitive power supplies. Each supply consists of 24 capacitor banks, a protection inductor, crowbar circuit, dump circuit. To simplify system configuration, two supplies are equipped with one charge unit, one thyristor switch and one switchgear. The magnet wound with CuNb wire reinforced internally with Zylon fiber composite and externally with stainless steel shell is designed to provide 60T/60ms magnetic field in a bore of 26mm. The control system provide integrated charge/discharge procedure and a remote control terminal. The whole system will be installed in May 2017, and the Preliminary test results will provided in the paper.

Acknowledgements: The authors would like to acknowledge the supports of the National key research and development program of China (2016YFA0401702) and the Program for New Century Excellent Talents in University.

Submitters Country

China

Primary authors: JIANG, Chengxi (Huazhong University of Science and Technology); DING, Hongfa (Wuhan National High Magnetic Field Center, Huazhong University of Science and Technology); Prof. LI, Liang (Huazhong University of Science and Technology); HAN, Xiaotao (Wuhan National High Magnetic Field Center, Huazhong University of Science and Technology); PENG, Tao (Wuhan National High Magnetic Field Center)

Presenter: JIANG, Chengxi (Huazhong University of Science and Technology)

Session Classification: Mon-Af-Po1.12

Track Classification: H2 - Power Supplies and Flux Pumps