



Contribution ID: 1297  
 competition)

Type: Oral (Student, Not in Competition) / Orale (Étudiant(e), pas dans la

## Effect of temperature on the generation of acoustic harmonics in a tissue-mimicking liquid

Wednesday, 15 June 2016 13:30 (15 minutes)

In this study, the temperature dependence of acoustic harmonics generated by nonlinear ultrasound beam propagation in a mixture of 90% glycerol and 10% water (by volume) used as a tissue-mimicking liquid, was studied as a function of temperature from 26°C to 46°C at two different frequencies. Simulations were performed with a temperature-dependent Khokhlov–Zabolotskaya–Kuznetsov (KZK) nonlinear acoustic beam propagation model. Two single-element focused transducers with center frequencies of 5 and 13 MHz were used to transmit high pressure ultrasound signals. The ultrasound signals along the acoustic axis of each transducer were recorded using a needle hydrophone. The axial components of the fundamental (p1), the second (p2) and the third (p3) harmonics were calculated from the measured and simulated signals at temperatures of 26, 31, 36, 41 and 46°C for both transmit frequencies. At the transmit frequency of 5 MHz, the peak values of the harmonics p1, p2, and p3, and their ratios p2/p1, p3/p1 increased by 35%, 189%, 573%, 113% and 396%, respectively in simulations and by 35% ± 2.5%, 223% ± 17%, 797% ± 38%, 138% ± 9% and 576% ± 31%, respectively in measurements as the temperature elevated from 26°C to 46°C. The peak values of p1, p2, and p2/p1 increased by 256%, 1773% and 426%, respectively in simulations and by 205% ± 15%, 1411% ± 103%, and 394% ± 17%, respectively in measurements at 13 MHz for the same temperature range. Moreover, the p3 and p3/p1 increased by 597% and 311%, respectively in simulations and by 387% ± 31% and 204% ± 37%, respectively in measurements at 13 MHz as the temperature was increased from 36 to 46°C. The significant increase in the rate of generation of harmonics with temperature is due to the combined effects of increase in the nonlinearity parameter (B/A) and reduction in both the attenuation coefficient and the speed of sound with temperature in the medium.

**Primary author:** Mr MARAGHECHI, Borna (Department of Physics, Ryerson University, 350 Victoria Street, Toronto, Ontario, M5B 2K3, Canada)

**Co-authors:** Dr TAVAKKOLI, Jahan (Department of Physics, Ryerson University, 350 Victoria Street, Toronto, Ontario, M5B 2K3, Canada. Institute for Biomedical Engineering, Science and Technology (iBEST), Keenan Research Centre for Biomedical Science, St. Michael's Hospital, Toronto, Ontario, M5B 1W8, Canada); Prof. KOLIOS, Michael C. (Department of Physics, Ryerson University, 350 Victoria Street, Toronto, Ontario, M5B 2K3, Canada. Institute for Biomedical Engineering, Science and Technology (iBEST), Keenan Research Centre for Biomedical Science, St. Michael's Hospital, Toronto, Ontario, M5B 1W8, Canada)

**Presenters:** Mr MARAGHECHI, Borna (Department of Physics, Ryerson University, 350 Victoria Street, Toronto, Ontario, M5B 2K3, Canada); Mr SHASWARY, Elyas (Dept. of Physics, Ryerson University)

**Session Classification:** W2-1 Biomechanics and Fluid Dynamics (DPMB) / Biomécanique et dynamique des fluides (DPMB)

**Track Classification:** Physics in Medicine and Biology / Physique en médecine et en biologie (DPMB-DPMB)