Advanced Detectors for Nuclear, High Energy and Astroparticle Physics

Contribution ID: 23

Type: Poster

Some aspects of characterization of GEM detector

Thursday 16 February 2017 18:20 (5 minutes)

We have studied the variation of the gain of a triple GEM detector as a function of the relative humidity. The temperature/pressure (T/p) corrected normalized gain has been found to decrease with decrease in relative humidity. The uniformity of gain over the entire surface of the detector has also been studied measuring the anode current. The details of the measurement process and the experimental results will be presented.

Presentation type

Oral

Authors: NAG, Dipanjan (Bose Institute); Dr BISWAS, Saikat (Bose Institute, 93/1 APC Road, Kolkata, INDIA); DAS, Supriya (Bose Institute); GHOSH, Sanjay Kumar (Bose Institute (IN)); Mr PAUL, Dipankar; PRASAD, Sidharth Kumar (Bose Institute (IN)); Mr RAHA, Sibaji; ROY, SHREYA (BOSE INSTITUTE)

Presenters: NAG, Dipanjan (Bose Institute); Dr BISWAS, Saikat (Bose Institute, 93/1 APC Road, Kolkata, INDIA); DAS, Supriya (Bose Institute); GHOSH, Sanjay Kumar (Bose Institute (IN)); Mr PAUL, Dipankar; PRASAD, Sidharth Kumar (Bose Institute (IN)); Mr RAHA, Sibaji; ROY, SHREYA (BOSE INSTITUTE)

Session Classification: Poster session