

CMS ECAL DAQ system

Tuesday, July 10, 2018 4:40 PM (20 minutes)

In 2017 the Large Hadron Collider (LHC) at CERN has provided an astonishing 50 fb⁻¹ of proton-proton collisions at a center of mass energy of 13 TeV. The Compact Muon Solenoid (CMS) detector has been able to record the 90.3% of this data. During this period, the CMS Electromagnetic Calorimeter (ECAL), based on 75000 scintillating PbWO₄ crystals and a silicon and lead preshower, has continued exhibiting excellent performance with a very stable data acquisition (DAQ) system. The ECAL DAQ system follows a modular and scalar schema: the 75000 crystals are divided in sectors (FED), each of them controlled by 3 interconnected boards. These board are responsible for the configuration and control of the front-end electronics configuration, the generation of trigger primitive for the central CMS L1 trigger, and for the collection of data. A multi-machine distributed software configures the electronic boards and follows the life cycle of the acquisition process. The ECAL electronic modular configuration is reflected in the software where a tree control structure is applied. Through a master web application, the user controls the communication with the sub-applications which are responsible for the off-detector board configurations. Since the beginning of Run 2 in 2015, many improvements to the ECAL DAQ have been implemented to reduce occasional errors, as well as to mitigate single event upsets in the front-end electronics, and to improve the efficiency. Efforts at the software level have been made to introduce automatic recovery in case of errors. These procedures are mandatory to have a reliable and efficient acquisition system.

Primary author: CUCCIATI, Giacomo (CERN)

Presenter: CUCCIATI, Giacomo (CERN)

Session Classification: Posters

Track Classification: Track 1 - Online computing