

Gas Analysis Update

- RICH1 OK
 Air quantity increasing
- RICH2 OK
 Three different calibs ?
- MWPC PROBLEM
 Concentration values
 too far from ideal ones

LHCb

- Bronkhorst pressure
 regulator installed
- Problems arisen after its configuration
- => No reliable results for analysis done within the last two weeks





CMS

Bronkhorst

GC Tracker Status

Experimental difficulties



- C6F14 low vapour pressure
- Liquefies very easily (at room temperature)
- Mixture test with different ratios
 - No C6F14 peak in the chromatograms
 - C6F14 Appeared only in this particular condition:
 - N2 flow = 4 l/h, C6F14 flow = 1 l/h
 - Heating band turned on
 - 10 minutes delay between each analysis
 - I wasn't able to replicate the result later

Removing the filter can be dangerous for the GC



