



Contribution ID: 288

Type: **Contributed**

## Combining NEG pump and an XHV BNNT cryopump

*Tuesday 19 June 2018 10:10 (20 minutes)*

We have assembled a system using an array of NEG modules and a cryopump with Boron Nitride Nanotubes (BNNT) instead of the traditional charcoal. The BNNT has been mechanically attached to the cryosorption surfaces of a commercial cryopump, and the system fully baked to remove water with no adhesive present in the system. We report here on the pump speed of the BNNT cryopump, and characterize the base pressure achieved in the combined NEG/cryopump system using both an extractor gauge and a Watanabe 3BG XHV ionization gauge.

**Author:** STUTZMAN, Marcy (Jefferson Lab)

**Co-authors:** Ms SEGOVIA MIRANDA, Anahi (Universidad Autonoma de Zacatecas); Mr ADDERLEY, Philip (Jefferson Lab); Dr POELKER, Matthew (Jefferson Lab)

**Presenter:** STUTZMAN, Marcy (Jefferson Lab)

**Session Classification:** Vacuum Science & Technology

**Track Classification:** Vacuum Science & Technology