

Study on the noise temperature of Josephson Parametric Amplifier (JPA) used in the axion dark matter search experiment at CAPP/IBS in KAIST

Friday 6 July 2018 20:15 (15 minutes)

In an axion dark matter search haloscope experiment, the noise temperature of a linear amplifier is a crucial component that seriously affects the sensitivity of the dark matter axion search. In the CAPP18T experiment at CAPP/IBS in KAIST, we use a cryogenically cooled Josephson Parametric Amplifier (JPA) in order to amplify weak RF signals from a resonant cavity. In this presentation, we describe a method to obtain an accurate and repeatable input noise temperature of the JPA: We use a device with a cryogenic attenuator co-located with the amplifier. A dilution refrigerator ($\sim 20\text{mK}$) and a cryogenic High Electron Mobility Transistor are used for this measurement. We will also discuss the calibration techniques.

Authors: LEE, Ji-Young (Institute for Basic Science); Dr MIN, Byeonghun (IBS); Dr KIM, DongLak (IBS); Dr KIM, JinGeun (IBS); Mr PARK, HeeJun (IBS); Mr KIM, Jongkuk (Korea Advanced Institute for Science and Technology); Mr LEE, YoungJae (KAIST); Mr AHN, Moohyun (Seoul National University); YOO, JONGHEE (KAIST)

Presenter: LEE, Ji-Young (Institute for Basic Science)

Session Classification: POSTER

Track Classification: Dark Matter Detection