



Contribution ID: 619

Type: **Contributed Oral Presentation**

## **C2Or2A-05: Investigation of improving cool-down speed of Stirling type pulse tube cryocooler with ambient displacer**

*Tuesday, July 23, 2019 4:30 PM (15 minutes)*

For a good cryocooler, besides high efficiency at working temperature, the cool-down speed is also important for some applications. Because of the difference in structure and operation mechanism, different cryocoolers have different characteristics of cool-down speed. This work introduces a Stirling type pulse tube cryocooler with ambient displacer which works at liquid nitrogen temperature region, and a strategy for improving the cool-down speed from room temperature to liquid nitrogen temperature has been investigated through simulation. The pulse tube cryocooler can provide 17.8 W cooling power at 77 K with 180 W input acoustic power at 70 Hz. By fixing the maximum displacer movement in the simulation, we investigate how tuning frequency around 70 Hz can change the cool-down speed. The strategy turns out to be effective in the beginning stage of the cool-down process, an improvement of speed by about 15-20%. The gain becomes less obvious as the cold head temperature drops. In conclusion, tuning frequency during cool-down process helps the cold-head to reach final temperature faster but the average gain is not as big as expected.

**Primary authors:** YAN, Cungang (Technical Institute of Physics and Chemistry, Chinese Academy of Sciences ; University of Chinese Academy of Sciences); Prof. WANG, Xiaotao (Technical Institute of Physics and Chemistry, Chinese Academy of Sciences); Prof. DAI, Wei (Technical Institute of Physics and Chemistry, Chinese Academy of Sciences); Prof. LUO, Ercang (Technical Institute of Physics and Chemistry, Chinese Academy of Sciences)

**Presenters:** YAN, Cungang (Technical Institute of Physics and Chemistry, Chinese Academy of Sciences ; University of Chinese Academy of Sciences); Prof. WANG, Xiaotao (Technical Institute of Physics and Chemistry, Chinese Academy of Sciences); Prof. DAI, Wei (Technical Institute of Physics and Chemistry, Chinese Academy of Sciences)

**Session Classification:** C2Or2A - Non-Aerospace Stirling and Pulse Tube Cryocoolers