Contribution ID: 1709

**Type: Contributed Oral Presentation** 

## Thu-Af-Or24-03: Novel Characterization Technique to Visualize Local Defects in a REBCO Pancake Coil Winding

Thursday 26 September 2019 17:15 (15 minutes)

It is relevant to establish a reliable coil winding techniques which is free from local defects and/or delamination of REBCO tape strand. A coil is usually tested by measuring transport current-voltage characteristics as a whole, however, the position and distribution of the defects can not be identified from such global measurements. This prevent us from clarifying the mechanism of the degladation by the coiling. In this study, we have succeeded in developing a novel method to visualize local defects in a pancake coil winding with a high spatial resolution of a single tape thickness. This allows us to collect information on local degradation in the coil winding, i.e., a powerful method as a diagnostic technique in order to screening out a degraded coil. This is also very useful to investigate degradation mechanism and to improve coil winding methods. We adopted this method to investigate the properties of a large shielding coil used for a half-size REBCO based MRI magnet. Test results using an artificial tape stack including a single non-superconducting tape with a thickness of 100 um and the measurement results on the actual sheilding coil will be presented.

This work was supported by the New Energy and Industrial Technology Development Organization (NEDO).

**Authors:** KISS, Takanobu (Kyushu University); HIGASHIKAWA, Kohei (Kyushu University); Mr OHTA, Shogo (Kyushu University); Mr TAKARABE, Yusuke (Kyushu Universit); Mr IMAMURA, Kazutaka (Kyushu University); Mr MIURA, Hideaki (Mitsubishi Electric Croporation); YOKOYAMA, Shoichi (Mitsubishi Electric Croporation)

Presenter: KISS, Takanobu (Kyushu University)

Session Classification: Thu-Af-Or24 - Diagnostics and Test Results of Coils