



Contribution ID: 1213

Type: **Poster Presentation**

Tue-Mo-Po2.07-03 [43]: Charging, Discharging, and Over-current Characteristics of Partially Insulated MgB₂ Magnet Using Cr-coated MgB₂ Wires

Tuesday 24 September 2019 08:45 (2 hours)

Recently, we developed a 0.5-T/300 mm MgB₂ magnet using the partial insulation (PI) winding technique that only employs layer-to-layer insulations without the turn-to-turn insulations; the magnet was self-protective but has a significant charge-discharge delay. In this study, the use of Cr-coated MgB₂ wires was proposed as an alternative solution to enhance the characteristic resistance of the MgB₂ magnet, in combination with the PI winding technique. The charging-discharging tests confirmed that the charging-discharging rates of the PI magnet that uses Cr-coated wires was ten times faster than the magnet that utilizes uncoated wires. In addition, the self-protective behavior of the PI magnet that uses the Cr-coated wires were verified via the over-current tests. Herein, the feasibility of employing the proposed winding technique for the development of self-protective MgB₂ magnetic resonance imaging magnets with fast charging-discharging rates was discussed in detail.

[Acknowledgement]

This work was supported by the Materials and Components Technology Development Program of KEIT [10053590, Development of MgB₂ wire and coil with a high critical current and long length for superconducting medical electric power equipment].

Primary authors: Mr KIM, Young-Gyun (Korea University); LEE, Haigun (Korea University)

Co-authors: KIM, Jiman (Kiswire Advanced Technology); Mr YOO, Byeongha (Korea University); Mr KIM, Kihyun (Korea University); Mr HWANG, Duck Young (Kiswire Advanced Technology Co); KIM, Ji Hyung (Jeju National University); KIM, Ho Min (Jeju National University)

Presenter: Mr KIM, Kihyun (Korea University)

Session Classification: Tue-Mo-Po2.07 - MRI Magnets III