

VNA measurements of 6mm buttons in button block

Tuesday 4 June 2019 13:55 (15 minutes)

The S-parameters of a four button block with 6mm buttons have been measured using a commercial Vector Network Analyser. The quality of the measurement is evaluated with the Lambertson method in mind to determine the gain of the individual buttons.

Author: Dr REHM, Guenther (Diamond Light Source)

Presenter: Dr REHM, Guenther (Diamond Light Source)

Track Classification: BPM error compensation schemes