

# Advancing the JUNGFRÄU detector towards low-energy X-ray applications

Monday, September 13, 2021 10:30 AM (15 minutes)

The charge integrating hybrid silicon pixel detector JUNGFRÄU has found widespread use for hard X-ray applications at free-electron laser (FEL) and synchrotron facilities. Equipped with three dynamic switching gains at pixel level, the most recent version of JUNGFRÄU offers single photon resolution from at least 1.2 keV and a high dynamic range of  $10^4$  12 keV photons at a maximum frame rate of 2 kHz. The detector hardware is modular with a single module comprising 0.5 megapixels at a pixel size of  $75 \times 75 \mu\text{m}^2$ . Owing to the low noise performance, high dynamic range, position resolution, and easy scalability of the JUNGFRÄU system, the detector is of high interest for applications in soft X-ray science.

Recently, a four-megapixel (4M) JUNGFRÄU camera has been installed at the Maloja end station of the low energy beamline Athos of the Swiss free-electron laser (SwissFEL). The beamline operates at a photon energy range from 250 eV to 2 keV, making the current JUNGFRÄU system applicable for experiments at the higher end of the available energy spectrum. At energies below  $\sim 1$  keV, the readout capacitance of the hybrid detector limits the capability of resolving single photons. Signal amplification at sensor level could overcome this limitation and allow the JUNGFRÄU system to significantly extend its low energy range.

In this contribution, we present first results of the Maloja JUNGFRÄU 4M system, demonstrating the capabilities of the current detector version for soft X-ray science, and provide an outlook on efforts to couple the JUNGFRÄU readout ASIC with LGAD sensors to achieve single photon resolution at energies below 1 keV.

## Title

## Your name

Viktoria Hinger

## Institute

Paul Scherrer Institut

## email

viktoria.hinger@psi.ch

## Nationality

**Primary authors:** MOZZANICA, Aldo (PSI - Paul Scherrer Institut); Dr AL HADDAD, Andre (Paul Scherrer Institut); Dr SCHMITT, Bernd (Paul Scherrer Institut); OZEROV, Dmitry (Paul Scherrer Institut); Dr SCHNORR, Kirsten Andrea (Paul Scherrer Institut); Dr HINGER, Viktoria (Paul Scherrer Institut)

**Co-authors:** Dr BERGAMASCHI, Anna (PSI); LOPEZ CUENCA, Carlos (PSI - Paul Scherrer Institut); MEZZA, Davide (Paul Scherrer Institut); Dr GREIFFENBERG, Dominic (PSI - Paul Scherrer Institute); Dr FRÖJDH, Erik (PSI - Paul Scherrer Institute); Dr ZHANG, Jianguo (Paul Scherrer Institut); MOUSTAKAS, Konstantinos (PSI - Paul Scherrer Institut); BRÜCKNER, Martin (PSI - Paul Scherrer Institut); KOZŁOWSKI, Pawel (PSI - Paul Scherrer Institut); BARTEN, Rebecca (Paul Scherrer Institut); DINAPOLI, Roberto (Paul Scherrer Institut); Dr CHIRIOTTI

ALVAREZ, Sabina (PSI - Paul Scherrer Institut); HASANAJ, Shqipe (Paul Scherrer Institut); KING, Thomas (Paul Scherrer Institut)

**Presenter:** Dr HINGER, Viktoria (Paul Scherrer Institut)

**Session Classification:** Detectors for FELS, Synchrotrons and Other Advanced Light Sources

**Track Classification:** Detectors for FELS, Synchrotrons and Other Advanced Light Sources