

Panel TOF-PET imager

Thursday, September 16, 2021 10:43 AM (1 minute)

Positron emission tomography (PET) is one of the most important diagnostic tools in medicine, providing three-dimensional imaging of functional processes in the body. The method is based on detecting two gamma rays originating from the point of annihilation of the positron emitted being by radio-labelled agent, and used to follow the human's physiological processes. In Time-Of-Flight PET gamma rays' arrival time is measured in addition to their position. The coincidence timing resolution (CTR) of state-of-the art scanners is between 200 ps and 500 ps FWHM, which can already significantly improve the contrast in imaging large objects. To increase the sensitivity of the next-generation PET scanners timing accuracy should be substantially increased. By using latest advances multichannel system with improved CTR is becoming technologically possible. Generally 3D images from limited angle PET scanners are distorted and have artefacts. Fortunately, with improving timing resolution of PET gamma detectors, artefact free images can be obtained even by a very simplified detector. In the contribution we will show the simulation studies of the simple panel detector using gamma detectors with 50 ps coincidence timing resolution. With this new concept, the price of PET scanners for imaging single or multiple organs can be drastically decreased. We evaluated different panel detector arrangements by imaging different phantoms. We compared the reconstructed images with the image obtained with the Siemens Biograph Vision, state-of-the-art clinical PET scanner. We found comparable image quality parameters of both systems when the CTR approaches 50ps FWHM and also that good CTR can partially compensate for smaller gamma detection efficiency.

Title

Prof.

Your name

Rok Pestotnik

Institute

Jožef Stefan Institute

email

rok.pestotnik@ijs.si

Nationality

Slovene

Primary author: PESTOTNIK, Rok (Jozef Stefan Institute (SI))

Co-authors: DOLENEC, Rok (Institut "Jožef Stefan"); Prof. EL FAKHRI, Georges (Massachusetts General Hospital); Prof. MAJEWSKI, Stan (University of California Davis); STUDEN, Andrej (Jozef Stefan Institute (SI)); RAZDEVŠEK, Gašper (University of Ljubljana); KORPAR, Samo; KRIZAN, Peter (University of Ljubljana)

Presenter: PESTOTNIK, Rok (Jozef Stefan Institute (SI))

Session Classification: Poster Session 5 (Gas-based Detectors; Medical Applications of Position Sensitive Detectors)

Track Classification: Medical Applications of Position Sensitive Detectors