

Measurement of angular correlations in positronium decay using GaGG scintillator matrices with single-side readout by silicon photo-multipliers

Wednesday 15 September 2021 12:59 (1 minute)

Gamma-ray polarization is of prime interest in many areas of physics. One particular is biomedical imaging with positron emission tomography (PET). Two orthogonally polarized, entangled gamma-rays are emitted in an event of para-positronium annihilation thus they are strongly correlated. When they undergo Compton scattering, the initial correlation dominantly results in their orthogonal azimuthal scattering. This correlation can be used in PET as an additional criterium for distinguishing a true coincidence event from the background where such correlation is lacking. In a previous study we showed a moderate azimuthal correlation could be observed via Compton scattering in system of two 4x4 Lutetium Fine Silicate (LFS) scintillation matrices. The goal of the present study is to increase the detector sensitivity to azimuthal correlation, where one of the important parameters is the energy resolution. Hence, we measured azimuthal correlations of annihilation quanta with two 8x8 matrices of Gadolinium Aluminum Gallium Garnet doped with Cerium (GaGG:Ce). Each detector matrix contains 64 crystals of size 3 mm x 3 mm x 20 mm and it is read-out by a single Silicon Photo-multiplier (SiPM) array, with one SiPM matching one pixel. The studied single-side readout concept keeps the modules compact and cost-efficient on a large scale. Coincidence events were recorded using a Na-22 source placed between the modules and the mean energy resolution at 511 keV was $9.7\% \pm 0.5\%$ and $11.3\% \pm 0.7\%$ for the two modules, respectively. We clearly identified and reconstructed the Compton scattering events and we observe the excess of orthogonally scattered gammas over the ones with the parallel azimuthal angles. We present the measured azimuthal modulation factors for several kinematic selection criteria and different inter-pixel distances and we compare them to previous findings with 4x4 LFS scintillators. Finally, we discuss the perspective of using the presented concept in PET.

Title

Dr

Your name

Siddharth Parashari

Institute

Department of Physics, Faculty of Science, University of Zagreb, Bijenička cesta 32, 10000, Zagreb

email

siddharthparashari5@gmail.com

Nationality

Indian

Primary authors: PARASHARI, Siddharth; Prof. BOSNAR, Damir (Faculty of Science); Ms KOŽULJEVIĆ, Ana Marija (University of Zagreb); MAKEK, Mihael (Faculty of Science, University of Zagreb)

Presenter: PARASHARI, Siddharth

Session Classification: Poster Session 2 (X-ray and Gamma Ray Detectors; Applications in Nuclear Physics and Nuclear Industry; Detectors for FELs, Synchrotrons and Other Advanced Light Sources; Detectors for Neutron Facilities; Novel Ionising Radiation Detection Systems)

Track Classification: X-ray and Gamma Ray Detectors