

Detection module based on position-sensitive large-area Silicon photomultipliers

Thursday, 16 September 2021 10:35 (1 minute)

Silicon Photomultipliers are compact single-photon-sensitive detectors, widely used in many applications. In FBK (Trento, Italy) we developed large area SiPMs (up to $10 \times 10 \text{ mm}^2$), based on different technologies and we are also focusing on the position-sensitive SiPM (PS-SiPM) technology based on charge-sharing approach. These are based on the so called “linearly-graded, LG” technology, exploiting a weighted current dividers on vertical and horizontal axis, to obtain signals (left, right, top and bottom signals) with amplitude and charge proportional to the light-absorption position.

Such large area detector with position sensitivity is very interesting in applications like ultra-high spatial resolution, MR-compatible PET and in the creation of a compact gamma and beta cameras with a reduced number of channels, for radio-guided surgery or other clinical decision support tool for diagnostic imaging. In this contribution, we illustrate the project developed in collaboration with the University of Geneva for a detection module based on a 2×2 tile of large-area PS-SiPMs, including front-end amplifiers and shaped like a “handable” probe. Total area is $1.6 \times 1.6 \text{ cm}^2$. The PS-SiPMs are connected in a smart configuration, maintaining a very low number of channels. Indeed, while for a single LG-SiPM there are 4 output, for a 2×2 tile, we have just 6 outputs. The measured position resolution (measured with a pulsed LED, scanned over active area) is better than 0.5 mm.

Finally, we performed an estimation of the performance of such module when used as a gamma camera for tumor detection. With OpenGate simulation, we proved that for a radio-tracer emitting gamma of lower energy, as $^{99\text{m}}\text{Tc}$, the gamma camera could achieve an excellent performance. Superficial tumors, of about 2 mm, could be well reconstructed in less than 20 seconds. Tumors bigger than 6 mm in radius could be detected at 30 mm deep within about 10 seconds.

Title

Dr

Your name

Fabio Acerbi

Institute

FBK

email

acerbi@fbk.eu

Nationality

Primary authors: Dr ACERBI, Fabio (Fondazione Bruno Kessler (FBK)); NAGAI, Andrii (Universite de Geneve (CH)); Dr MERZI, Stefano (FBK); VENTURA BARROSO, Ana (Universite de Geneve (CH)); DELLA VOLPE, Domenico (Universit  de Gen ve); Dr GOLLA, Alberto (Fondazione Bruno Kessler)

Presenter: Dr ACERBI, Fabio (Fondazione Bruno Kessler (FBK))

Session Classification: Poster Session 5 (Gas-based Detectors; Medical Applications of Position Sensitive Detectors)

Track Classification: Medical Applications of Position Sensitive Detectors