



Contribution ID: 285 Contribution code: TUE-PO1-115-01

Type: Poster

## A Low-Temperature-Dependent Calibration of Hall probes for CPMU

Tuesday, 16 November 2021 13:15 (20 minutes)

### *Abstract*

A cryogenic permanent-magnet undulator (CPMU) with a period of 18mm and a magnetic length of 2m is being constructed for the Taiwan Photon Source (TPS). The CPMU with a gap of 5.5 mm can generate an effective magnetic field of 1.2 T at 150 K. When the field measurement, the temperature of hall probe will be drop at cryogenic temperature. Therefore, a field strength calibration and thermostatic temperature control system is necessary. The range of field strength calibration is from 5.5 to  $\pm 1.5$ T at a homemade dipole electromagnet. A 2-axis compact SENIS hall probe is mounted on homemade copper plate in vacuum chamber. The temperature control system consists of a cryocooler, PT100 sensor and a heater to control the Hall probe temperature. Finally, a higher-order polynomial surface fitting to analysis measurement data from calibration system. The field strength maximum error is  $< 0.2$  G at fitting surface. The detailed temperature dependent calibration system is presented in this poster.

**Primary author:** Dr CHEN, Chih Wei (National Synchrotron Radiation Research Center)

**Co-authors:** HSU, Yang-Yang (National Synchrotron Radiation Research Center); YANG, Chin-Kang (National Synchrotron Radiation Research Center); CHEN, Hsiung (National Synchrotron Radiation Research Center); HUANG, Jui-Che (National Synchrotron Radiation Research Center)

**Presenter:** Dr CHEN, Chih Wei (National Synchrotron Radiation Research Center)

**Session Classification:** TUE-PO1-115 Permanent accelerator magnets