



Canadian Association
of Physicists

Association canadienne
des physiciens et physiciennes

Contribution ID: 468

Type: Oral (Non-Student) / Orale (non-étudiant(e))

Cryogenic detector monitoring and calibration with internally mounted LEDs

Wednesday 9 June 2021 13:15 (10 minutes)

As dark matter searches aim to achieve lower energy thresholds, it is important to understand the behaviour of the detectors in these new regimes. Light-emitting diodes (LEDs) offer a simple and flexible source of photons with energy ranges from 0.3 eV (mid-infrared) to 5 eV (near ultraviolet). Prototype cryogenic silicon detectors developed by the SuperCDMS collaboration have been able to achieve energy resolutions below 3 eV. By taking advantage of the Neganov-Trofimov-Luke (NTL) effect, the effective resolution can be reduced to well below the bandgap of silicon at mK temperatures of 1.2 eV. At our R&D facility at TRIUMF, we have been testing two of these “HVeV” detectors using a range of LEDs at mK temperatures. We have demonstrated the LEDs suitability for detector calibration and have applied them to investigate the detector response to photons ranging from sub-gap infrared to near ultraviolet.

Primary authors: Dr MAYER, Adam (TRIUMF); FASCIONE, Eleanor (TRIUMF/Queen’s University); Dr GHAIH, Muad (Queen’s University); Dr RAU, Wolfgang (TRIUMF)

Presenter: Dr MAYER, Adam (TRIUMF)

Session Classification: W2-11 Test Facility I (PPD) / Installation pour tests I (PPD)

Track Classification: Particle Physics / Physique des particules (PPD)